

CTSU Lab Room Freezer Min/Max Temperature Log

Maintain temperature between -20°C to -30°C

-20 Accucold Freezer - Delp

2021 12 000719

Refrigerator or Freezer ID/Location:
CTSU RAINBOW

Month 9 September Year 2025 Threshold: 4 hours

Document the details of any temperature excursion in comments below. Alert study team if excursion exceeds threshold.

Date (Weekdays only)	Time	Actual Temp	Min °C	Max °C	Initials	Comments
Mon 9/1/2025					m	Closed Holiday
Tue 9/2/2025	0726	-22.5	-29.3	-22.4	nw	
Wed 9/3/2025	0857	-22.8	-29.1	-22.4	MW	
Thu 9/4/2025	0736	-24.3	-29.2	-22.5	MW	
Fri 9/5/2025	1315	-24.5	-29.4	-22.5	MW	
Mon 9/8/2025	1600	-24.4	-29.3	-22.5	ADK	
Tue 9/9/2025	0752	-25.3	-29.3	-22.5	MW	
Wed 9/10/2025	0715	-25.5	-28.4	-13.6	MW	
Thu 9/11/2025	0808	-25.5	-28.4	-13.6	MW	
Fri 9/12/2025	0800	-25.8	-29.6	-17.9	ADK	
Mon 9/15/2025	0830	-26	-29.3	-22.8	ADK	
Tue 9/16/2025	0815	-24.6	-28.4	-23.3	MW	
Wed 9/17/2025	0915	-27.8	-28.4	-19.9	MW	
Thu 9/18/2025	0945	-27.6	-28.6	-21.9	MW	
Fri 9/19/2025	0853	-24.2	-28.5	-23.2	MW	
Mon 9/22/2025	0800	-24.6	-28.3	-23.1	ADK	
Tue 9/23/2025	0910	-23.8	-28.7	-23.2	MW	
Wed 9/24/2025	0901	-24.1	-28.3	-23.1	MW	
Thu 9/25/2025	0754	-24.9	-28.4	-17.2	MW	
Fri 9/26/2025	0800	-23.6	-28.5	-23.0	ADK	
Mon 9/29/2025	0914	-26.7	-28.7	-23.1	MW	
Tue 9/30/2025	0754	-25.6	-28.6	-23.0	MW	
Wed 10/1/2025					N/A	

Key: * Heavy usage