

CTSU Lab Room Freezer Min/Max Temperature Log

Maintain temperature between 2°C to 8°C

Refrigerator- Sudler

ABS9576207-1907

Refrigerator or Freezer ID/Location:
CTSU RAINBOW

Month 10 October Year 2025 Threshold: 30 minutes

Document the details of any temperature excursion in comments below. Alert study team if excursion exceeds threshold.

	Date (Weekdays only)	Time	Actual Temp	Min °C	Max °C	Initials	Comments
Wed	10/1/2025	0722	4.5	2.4	5.7	MW	
Thu	10/2/2025	0800	3.3	2.0	6.0	MW	
Fri	10/3/2025	0755	3.3	1.7	8.2	MW	
Mon	10/6/2025	0800	4.8	2.7	5.7	ADK	
Tue	10/7/2025	0800	3.1	1.9	7.4	ADK	
Wed	10/8/2025	0715	4.6	2.0	8.0	MW	
Thu	10/9/2025	0800	2.3	4.0	10.0	MW	
Fri	10/10/2025	0720	2.6	6.0	12.0	MW	
Mon	10/13/2025	0756	5.5	2.5	9.7	MW	
Tue	10/14/2025	0715	5.1	2.2	6.0	MW	
Wed	10/15/2025	0755	3.5	2.1	6.1	MW	
Thu	10/16/2025	0715	3.5	2.1	6.6	MW	
Fri	10/17/2025	0715	4.4	1.5	5.9	MW	
Mon	10/20/2025	0718	4.4	1.5	5.9	MW	
Tue	10/21/2025	0714	4.4	1.5	5.9	MW	
Wed	10/22/2025	0735	5.1	1.3	7.0	MW	
Thu	10/23/2025	0745	3.4	1.5	5.7	MW	
Fri	10/24/2025	0745	4.4	1.8	6.7	MW	
Mon	10/27/2025	0715	2.9	2.8	5.8	MW	
Tue	10/28/2025	0735	5.6	1.5	7.8	MW	
Wed	10/29/2025	0715	4.4	1.6	7.6	MW	
Thu	10/30/2025	1000	5.2	2.0	6.3	MW	
Fri	10/31/2025	0730	5.2	2.5	5.9	MW	

Key: * Heavy usage