

**COBB Refrigeration & Laboratory Services LLC**

8071 Flint St, Lenexa, Kansas 66214

Phone 913 403-0088

Fax 913 403-0089

**CERTIFICATION OF ACCURACY**

**Customer Information**

**Company Name: KU Med Center CTSU**

**Company Address: 3901 Rainbow, Kansas City, KS**

**Contact Person/ Phone Number: Danielle Hoage 913-574-3059**

**Thermometer Information**

**Manufacturer: Covidien**

**Model: Filac EZ**

**Serial No: A1952323X**

**Department: CTSU**

	High	Medium	Low
<b>Certified Temperature (°C):</b>	41.5	38.9	36.9
<b>Observed Temperature (°C):</b>	40.8	38.8	37.1
<b>Observed Accuracy (°C):</b>	0.7	0.1	0.2

**Passed/Failed: Pass**

**Certification Date: 1/10/2024**

**Certification Due: 1/2025**

**Certifying Equipment**

**Manufacturer: Fluke**

**Model: 52II**

**Serial No: 15100368**

**Due Date: 2/1/2024**

**Service Technician: Derek Hulett**  
**(Print)**

**Date: 1/10/2024**

**Service Technician:**  
**(Sign)**

