

KU SCHOOL OF MEDICINE

INSPECTION SUMMARY - JANUARY 2022

<u>Control No.</u>	<u>WO No.</u>	<u>Description</u>	<u>Location</u>	<u>CompletionComments</u>
23487	342665	CENTRIFUGE/TABLETOP	LABORATORY	(TC 01/28/2022 02:54:33 PM, KING, AARON) PM INSPECTION PASSED
23498	342666	LIGHT/EXAMINATION/EYE/ENT	CLINIC	(TC 01/28/2022 02:55:07 PM, KING, AARON) PM INSPECTION PASSED
24554	342681	ASPIRATOR/TRACHEAL	CLINIC	(TC 01/28/2022 02:55:29 PM, KING, AARON) PM INSPECTION PASSED
28141	342862	OXIMETER/PULSE	EQUIPMENT ROOM	(TC 01/28/2022 03:02:12 PM, KING, AARON) PM INSPECTION PASSED
29309	342877	FREEZER/LABORATORY	LABORATORY	(TC 01/28/2022 02:56:21 PM, KING, AARON) PM INSPECTION PASSED
29915	342953	CENTRIFUGE/TABLETOP	LABORATORY	(TC 01/28/2022 02:56:50 PM, KING, AARON) PM INSPECTION PASSED
30825	342979	FREEZER/GENERAL	LABORATORY	(TC 01/28/2022 03:03:02 PM, KING, AARON) WO 346905 STARTED, BATTERY. REPAIR COMPLETED. PM INSPECTION PASSED
31880	343034	ELECTROCARDIOGRAPH	CLINIC	(TC 01/28/2022 02:57:40 PM, KING, AARON) PM INSPECTION PASSED
36016	343111	FREEZER/LABORATORY	LABORATORY	(TC 01/28/2022 02:58:03 PM, KING, AARON) PM INSPECTION PASSED
36074	343112	MONITOR/PHYSIOLOGICAL	CLINIC	(TC 01/28/2022 02:58:29 PM, KING, AARON) PM INSPECTION PASSED
37324	343213	MONITOR/PHYSIOLOGICAL	CLINIC	(TC 01/28/2022 02:58:53 PM, KING, AARON) PM INSPECTION PASSED
37327	343214	SIGNAL EXTRACTION/OXIMETER/PULSE	CLINIC	(TC 01/28/2022 03:05:19 PM, KING, AARON) COULD NOT LOCATE FOR INSPECTION
37328	343215	OXIMETER/PULSE	CLINIC	(TC 01/28/2022 03:06:00 PM, KING, AARON) COULD NOT LOCATE FOR INSPECTION
5164	343474	CURRENT CUSTOMER BOOK	ADMINISTRATOR	
41839	343696	FREEZER/LABORATORY	LABORATORY	(TC 01/28/2022 03:04:13 PM, KING, AARON) WO 319369 STARTED, BATTERY. REPAIR COMPLETED. PM INSPECTION PASSED
42531	343764	PUMP/INFUSION	CLINIC	(TC 01/28/2022 02:59:23 PM, KING, AARON) PM INSPECTION PASSED

KU SCHOOL OF MEDICINE INSPECTION SUMMARY - JANUARY 2022

<u>Control No.</u>	<u>WO No.</u>	<u>Description</u>	<u>Location</u>	<u>CompletionComments</u>
42635	343808	SCALE/FLOOR	CLINIC	(TC 01/28/2022 02:59:44 PM, KING, AARON) PM INSPECTION PASSED
43418	343910	REFRIGERATOR/GENERAL	LABORATORY	(TC 01/28/2022 03:00:16 PM, KING, AARON) PM INSPECTION PASSED
44240	343912	OTO/OPHTHALMOSCOPE	CLINIC	(TC 01/28/2022 03:00:44 PM, KING, AARON) PM INSPECTION PASSED