

Sudler 2012 CTSU Drug Fridge
for
Sudler 2012 CTSU Drug Fridge

Jan 01, 2024 - Feb 01, 2024

Author: Miranda Wright
mwright15@kumc.edu

Review & approval records location:

Summary Report "Sudler 2012 CTSU Drug Fridge" for Sudler 2012 CTSU Drug Fridge

Jan 01, 2024 - Feb 01, 2024

MACHINE DETAILS


Serial Number: **A434F1FF2A93**
Location: **Outside Lab**
Description: **A434F1FF2A93**
Installation date:

EQUIPMENT DETAILS

Category: **+4 Refrigerator**
Manufacturer:
Model:
Service Entered Date:

TEMPERATURE

AVG: **-0.38 °C** | MIN: **-6.79 °C** | MAX: **19.1 °C**



DATA LOGS

CENTRAL TIME (US & CANADA)	TEMPERATURE (°C)
2023-12-31 6:00 AM	-4.72 °C
2024-01-01 6:00 AM	-5.51 °C
2024-01-02 6:00 AM	-4.48 °C
2024-01-03 6:00 AM	-3.82 °C
2024-01-04 6:00 AM	-3.43 °C
2024-01-05 6:00 AM	-3.52 °C
2024-01-06 6:00 AM	-4.17 °C
2024-01-07 6:00 AM	-5.47 °C
2024-01-08 6:00 AM	-3.89 °C
2024-01-09 6:00 AM	-3.35 °C
2024-01-10 6:00 AM	-3.4 °C
2024-01-11 6:00 AM	-3.09 °C
2024-01-12 6:00 AM	-3.07 °C
2024-01-13 6:00 AM	18.9 °C
2024-01-14 6:00 AM	17.53 °C
2024-01-15 6:00 AM	15.9 °C
2024-01-16 6:00 AM	14.01 °C
2024-01-17 6:00 AM	-5.59 °C
2024-01-18 6:00 AM	-2.84 °C
2024-01-19 6:00 AM	-1.69 °C
2024-01-20 6:00 AM	-2.31 °C
2024-01-21 6:00 AM	-4.82 °C
2024-01-22 6:00 AM	-6.69 °C
2024-01-23 6:00 AM	-2.73 °C
2024-01-24 6:00 AM	-0.84 °C
2024-01-25 6:00 AM	-0.66 °C
2024-01-26 6:00 AM	-0.51 °C
2024-01-27 6:00 AM	-1.26 °C
2024-01-28 6:00 AM	-3.13 °C
2024-01-29 6:00 AM	-1.23 °C
2024-01-30 6:00 AM	-0.56 °C
2024-01-31 6:00 AM	-0.37 °C

THRESHOLD ALERTS

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
2024-01-16 10:51 AM	Sudler 2012 CTSU Drug Fridge	Temperature	Copy of Threshold Alert for Element-T 4	9.13 °C	10.0/-100.0 °C	Alert Cleared
●	2024-01-16 10:51 AM	Notification:Email and SMS sent to Miranda Wright. Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 9.13 °C, within the limit of -100.0 °C to 10.0 °C. Sudler 2012 CTSU Drug Fridge is located in Outside Lab. Alert Rule:Copy of Threshold Alert for Element-T 4				
●	2024-01-16 10:51 AM	Message:Temperature Clear: 9.13 °C, within the limit of -100.0 °C to 10.0 °C Alert Rule:Copy of Threshold Alert for Element-T 4				
●	2024-01-12 6:00 PM	Notification:Email and SMS sent to Miranda Wright. Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 10.66 °C, still above the limit of 10.0 °C. Sudler 2012 CTSU Drug Fridge is located in Outside Lab. Alert Rule:Copy of Threshold Alert for Element-T 4				
●	2024-01-12 6:00 PM	Message:Temperature Alert: 10.66 °C, above the limit of 10.0 °C Alert Rule:Copy of Threshold Alert for Element-T 4				
●	2024-01-12 5:55 PM	Notification:Email and SMS sent to Miranda Wright. Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 10.66 °C, still above the limit of 10.0 °C. Sudler 2012 CTSU Drug Fridge is located in Outside Lab. Alert Rule:Copy of Threshold Alert for Element-T 4				
●	2024-01-12 5:55 PM	Message:Temperature Alert: 10.66 °C, above the limit of 10.0 °C Alert Rule:Copy of Threshold Alert for Element-T 4				
●	2024-01-12 5:50 PM	Notification:Email and SMS sent to Miranda Wright. Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 10.65 °C, above the limit of 10.0 °C. Sudler 2012 CTSU Drug Fridge is located in Outside Lab. Alert Rule:Copy of Threshold Alert for Element-T 4				
●	2024-01-12 5:50 PM	Message:Temperature Alert: 10.65 °C, above the limit of 10.0 °C Alert Rule:Copy of Threshold Alert for Element-T 4				

CONNECTION ALERTS

No Alert Events for selected date range.

POWER ALERTS

No Alert Events for selected date range.

ERROR ALERTS

No Alert Events for selected date range.

Sudler 2012 CTSU Drug Fridge
for
Sudler 2012 CTSU Drug Fridge

Feb 01, 2024 - Mar 01, 2024

Author: Miranda Wright
mwright15@kumc.edu

Review & approval records location:

Summary Report "Sudler 2012 CTSU Drug Fridge" for Sudler 2012 CTSU Drug Fridge

Feb 01, 2024 - Mar 01, 2024


MACHINE DETAILS

Serial Number: **A434F1FF2A93**
Location: **SUDLER 2012 CTSU OUTSIDE LAB**
Description: **A434F1FF2A93**
Installation date:

EQUIPMENT DETAILS

Category: **+4 Refrigerator**
Manufacturer:
Model:
Service Entered Date:

TEMPERATURE AVG: **-0.39 °C** | MIN: **-3.91 °C** | MAX: **9.13 °C**



DATA LOGS

CENTRAL TIME (US & CANADA)	TEMPERATURE (°C)
2024-01-31 6:00 AM	0.06 °C
2024-02-01 6:00 AM	0.37 °C
2024-02-02 6:00 AM	0.17 °C
2024-02-03 6:00 AM	-0.12 °C
2024-02-04 6:00 AM	-0.47 °C
2024-02-05 6:00 AM	-0.99 °C
2024-02-06 6:00 AM	-0.48 °C
2024-02-07 6:00 AM	-0.11 °C
2024-02-08 6:00 AM	0.06 °C
2024-02-09 6:00 AM	0.46 °C
2024-02-10 6:00 AM	-0.61 °C
2024-02-11 6:00 AM	-1.69 °C
2024-02-12 6:00 AM	-0.98 °C
2024-02-13 6:00 AM	-0.81 °C
2024-02-14 6:00 AM	-0.26 °C
2024-02-15 6:00 AM	7.03 °C
2024-02-16 6:00 AM	-0.37 °C
2024-02-17 6:00 AM	-2.38 °C
2024-02-18 6:00 AM	-3.8 °C
2024-02-19 6:00 AM	-1.69 °C
2024-02-20 6:00 AM	-0.76 °C
2024-02-21 6:00 AM	-0.41 °C
2024-02-22 6:00 AM	0.03 °C
2024-02-23 6:00 AM	0.04 °C
2024-02-24 6:00 AM	-0.62 °C
2024-02-25 6:00 AM	-0.6 °C
2024-02-26 6:00 AM	-0.16 °C
2024-02-27 6:00 AM	0.48 °C
2024-02-28 6:00 AM	-0.53 °C
2024-02-29 6:00 AM	-0.52 °C

THRESHOLD ALERTS

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
2024-02-15 12:07 PM	Sudler 2012 CTSU Drug Fridge	Temperature	Delp G040 +4	0.84 °C	8.0/1.0 °C	
						<p>Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>● 2024-02-15 12:07 PM Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.84 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4</p> <p>● 2024-02-15 12:07 PM Message:Temperature Alert: 0.84 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4</p> <p>● 2024-02-15 11:46 AM Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>● 2024-02-15 11:46 AM Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.97 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4</p> <p>● 2024-02-15 11:46 AM Message:Temperature Alert: 0.97 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4</p> <p>● 2024-02-15 11:25 AM Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>● 2024-02-15 11:25 AM Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.90 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4</p> <p>● 2024-02-15 11:25 AM Message:Temperature Alert: 0.90 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4</p>
2024-02-15 4:46 AM	Sudler 2012 CTSU Drug Fridge	Temperature	Delp G040 +4	1.98 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-02-15 4:46 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.98 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-02-15 4:46 AM					Message:Temperature Clear: 1.98 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-02-02 9:20 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.30 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in Outside Lab. Alert Rule:Delp G040 +4
	2024-02-02 9:20 AM					Message:Temperature Alert: 0.30 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-02-02 9:00 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.22 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in Outside Lab. Alert Rule:Delp G040 +4
	2024-02-02 9:00 AM					Message:Temperature Alert: 0.22 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-02-02 8:39 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.36 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in Outside Lab. Alert Rule:Delp G040 +4
	2024-02-02 8:39 AM					Message:Temperature Alert: 0.36 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
2024-02-02 9:22 AM	Sudler 2012 CTSU Drug Fridge	Temperature	Delp G040 -20	0.27 °C	-18.0/-30.0 °C	

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-02-02 9:22 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.27 °C, still above the limit of -5.0 °C. Sudler 2012 CTSU Drug Fridge is located in Outside Lab. Alert Rule:Delp G040 -20
						Message:Temperature Alert: 0.27 °C, above the limit of -5.0 °C Alert Rule:Delp G040 -20
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-02-02 9:01 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.20 °C, still above the limit of -5.0 °C. Sudler 2012 CTSU Drug Fridge is located in Outside Lab. Alert Rule:Delp G040 -20
						Message:Temperature Alert: 0.20 °C, above the limit of -5.0 °C Alert Rule:Delp G040 -20
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-02-02 8:40 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.28 °C, above the limit of -5.0 °C. Sudler 2012 CTSU Drug Fridge is located in Outside Lab. Alert Rule:Delp G040 -20
						Message:Temperature Alert: 0.28 °C, above the limit of -5.0 °C Alert Rule:Delp G040 -20

CONNECTION ALERTS

No Alert Events for selected date range.

POWER ALERTS

No Alert Events for selected date range.

ERROR ALERTS

No Alert Events for selected date range.

Sudler 2012 CTSU Drug Fridge
for
Sudler 2012 CTSU Drug Fridge

Mar 01, 2024 - Apr 01, 2024

Author: Miranda Wright
mwright15@kumc.edu

Review & approval records location:

Summary Report "Sudler 2012 CTSU Drug Fridge" for Sudler 2012 CTSU Drug Fridge

Mar 01, 2024 - Apr 01, 2024


MACHINE DETAILS

Serial Number: **A434F1FF2A93**
Location: **SUDLER 2012 CTSU OUTSIDE LAB**
Description: **A434F1FF2A93**
Installation date:

EQUIPMENT DETAILS

Category: **+4 Refrigerator**
Manufacturer:
Model:
Service Entered Date:

TEMPERATURE AVG: **0.57 °C** | MIN: **-2.12 °C** | MAX: **3.97 °C**



DATA LOGS

CENTRAL TIME (US & CANADA)	TEMPERATURE (°C)
2024-02-29 5:00 AM	-0.22° C
2024-03-01 5:00 AM	-0.55° C
2024-03-02 5:00 AM	-0.34° C
2024-03-03 5:00 AM	-0.32° C
2024-03-04 5:00 AM	0.37° C
2024-03-05 5:00 AM	0.2° C
2024-03-06 5:00 AM	0.07° C
2024-03-07 5:00 AM	0.1° C
2024-03-08 5:00 AM	0.28° C
2024-03-09 5:00 AM	-0.99° C
2024-03-10 5:00 AM	-1.83° C
2024-03-11 5:00 AM	-0.62° C
2024-03-12 5:00 AM	0.02° C
2024-03-13 5:00 AM	0.03° C
2024-03-14 5:00 AM	1.9° C
2024-03-15 5:00 AM	1.43° C
2024-03-16 5:00 AM	1.37° C
2024-03-17 5:00 AM	1.22° C
2024-03-18 5:00 AM	0.5° C
2024-03-19 5:00 AM	0.42° C
2024-03-20 5:00 AM	0.71° C
2024-03-21 5:00 AM	0.75° C
2024-03-22 5:00 AM	0.9° C
2024-03-23 5:00 AM	0.36° C
2024-03-24 5:00 AM	-0.63° C
2024-03-25 5:00 AM	0.26° C
2024-03-26 5:00 AM	0.88° C
2024-03-27 5:00 AM	0.3° C
2024-03-28 5:00 AM	0.67° C
2024-03-29 5:00 AM	0.99° C
2024-03-30 5:00 AM	1.43° C
2024-03-31 5:00 AM	1.71° C

THRESHOLD ALERTS

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
2024-03-29 9:20 AM	Sudler 2012 CTSU Drug Fridge	Temperature	Delp G040 +4	1.05 °C	8.0/1.0 °C	Alert Cleared
						<p>Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.05 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.</p> <p>Alert Rule:Delp G040 +4</p>
						<p>Message:Temperature Clear: 1.05 °C, within the limit of 1.0 °C to 8.0 °C</p> <p>Alert Rule:Delp G040 +4</p>
						<p>Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.98 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.</p> <p>Alert Rule:Delp G040 +4</p>
						<p>Message:Temperature Alert: 0.98 °C, below the limit of 1.0 °C</p> <p>Alert Rule:Delp G040 +4</p>
2024-03-29 8:34 AM	Sudler 2012 CTSU Drug Fridge	Temperature	Delp G040 +4	1.1 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-29 8:34 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.10 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
						Message:Temperature Clear: 1.10 °C, within the limit of 1.0 °C to 8.0 °C
	2024-03-29 8:34 AM					Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-29 2:24 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.00 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
						Message:Temperature Alert: 1.00 °C, below the limit of 1.0 °C
	2024-03-29 2:24 AM					Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-29 2:02 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.95 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
						Message:Temperature Alert: 0.95 °C, below the limit of 1.0 °C
	2024-03-29 2:02 AM					Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-29 1:41 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.98 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
						Message:Temperature Alert: 0.98 °C, below the limit of 1.0 °C
	2024-03-29 1:41 AM					Alert Rule:Delp G040 +4
2024-03-29 1:23 AM	Sudler 2012 CTSU Drug Fridge	Temperature	Delp G040 +4	1.05 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-28 9:48 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.05 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-03-28 9:48 AM					Message:Temperature Clear: 1.05 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-27 11:13 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.87 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-03-27 11:13 PM					Message:Temperature Alert: 0.87 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-27 10:53 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.91 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-03-27 10:53 PM					Message:Temperature Alert: 0.91 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-27 10:32 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.98 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-03-27 10:32 PM					Message:Temperature Alert: 0.98 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
2024-03-27 10:15 PM	Sudler 2012 CTSU Drug Fridge	Temperature	Delp G040 +4	1.02 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-27 10:15 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.02 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-03-27 10:15 PM					Message:Temperature Clear: 1.02 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-27 10:02 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.99 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-03-27 10:02 PM					Message:Temperature Alert: 0.99 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-27 9:39 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.98 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-03-27 9:39 PM					Message:Temperature Alert: 0.98 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
2024-03-27 12:14 PM	Sudler 2012 CTSU Drug Fridge	Temperature	Delp G040 +4	1.08 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-27 12:14 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.08 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
	2024-03-27 12:14 PM					Message:Temperature Clear: 1.08 °C, within the limit of 1.0 °C to 8.0 °C
						Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-26 10:01 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.98 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
	2024-03-26 10:01 PM					Message:Temperature Alert: 0.98 °C, below the limit of 1.0 °C
						Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-26 9:39 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.94 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
	2024-03-26 9:39 PM					Message:Temperature Alert: 0.94 °C, below the limit of 1.0 °C
						Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-26 9:19 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.98 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
	2024-03-26 9:19 PM					Message:Temperature Alert: 0.98 °C, below the limit of 1.0 °C
						Alert Rule:Delp G040 +4
2024-03-26 9:27 AM	Sudler 2012 CTSU Drug Fridge	Temperature	Delp G040 +4	1.05 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-26 9:27 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.05 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
						Message:Temperature Clear: 1.05 °C, within the limit of 1.0 °C to 8.0 °C
	2024-03-26 9:27 AM					Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-26 2:02 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.89 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
						Message:Temperature Alert: 0.89 °C, below the limit of 1.0 °C
	2024-03-26 2:02 AM					Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-26 1:42 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.00 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
						Message:Temperature Alert: 1.00 °C, below the limit of 1.0 °C
	2024-03-26 1:42 AM					Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-26 1:18 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.97 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
						Message:Temperature Alert: 0.97 °C, below the limit of 1.0 °C
	2024-03-26 1:18 AM					Alert Rule:Delp G040 +4

2024-03-25 10:35 AM Sudler 2012 CTSU Drug Fridge Temperature **Delp G040 +4** 1.05 °C 8.0/1.0 °C Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-25 10:35 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.05 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
	2024-03-25 10:35 AM					Message:Temperature Clear: 1.05 °C, within the limit of 1.0 °C to 8.0 °C
						Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-22 9:05 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.97 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
	2024-03-22 9:05 PM					Message:Temperature Alert: 0.97 °C, below the limit of 1.0 °C
						Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-22 8:42 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.99 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
	2024-03-22 8:42 PM					Message:Temperature Alert: 0.99 °C, below the limit of 1.0 °C
						Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-22 8:20 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.00 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
	2024-03-22 8:20 PM					Message:Temperature Alert: 1.00 °C, below the limit of 1.0 °C
						Alert Rule:Delp G040 +4
2024-03-22 8:20 AM	Sudler 2012 CTSU Drug Fridge	Temperature	Delp G040 +4	1.05 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-22 8:20 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.05 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-03-22 8:20 AM					Message:Temperature Clear: 1.05 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-22 6:45 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.91 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-03-22 6:45 AM					Message:Temperature Alert: 0.91 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-22 6:24 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.94 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-03-22 6:24 AM					Message:Temperature Alert: 0.94 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-22 6:04 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.95 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-03-22 6:04 AM					Message:Temperature Alert: 0.95 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4

2024-03-22 5:41 AM Sudler 2012 CTSU Drug Fridge Temperature **Delp G040 +4** 1.12 °C 8.0/1.0 °C Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-22 5:41 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.12 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
	2024-03-22 5:41 AM					Message:Temperature Clear: 1.12 °C, within the limit of 1.0 °C to 8.0 °C
						Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-22 3:38 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.97 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
	2024-03-22 3:38 AM					Message:Temperature Alert: 0.97 °C, below the limit of 1.0 °C
						Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-22 3:16 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.97 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
	2024-03-22 3:16 AM					Message:Temperature Alert: 0.97 °C, below the limit of 1.0 °C
						Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-22 2:56 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.98 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
	2024-03-22 2:56 AM					Message:Temperature Alert: 0.98 °C, below the limit of 1.0 °C
						Alert Rule:Delp G040 +4
2024-03-21 10:21 AM	Sudler 2012 CTSU Drug Fridge	Temperature	Delp G040 +4	1.04 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-21 10:21 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.04 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-03-21 10:21 AM					Message:Temperature Clear: 1.04 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-21 12:42 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.93 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-03-21 12:42 AM					Message:Temperature Alert: 0.93 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-21 12:21 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.00 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-03-21 12:21 AM					Message:Temperature Alert: 1.00 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-20 11:57 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.98 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-03-20 11:57 PM					Message:Temperature Alert: 0.98 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
2024-03-20 10:33 AM	Sudler 2012 CTSU Drug Fridge	Temperature	Delp G040 +4	1.07 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-20 10:33 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.07 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-03-20 10:33 AM					Message:Temperature Clear: 1.07 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-20 12:47 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.88 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-03-20 12:47 AM					Message:Temperature Alert: 0.88 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-20 12:25 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.97 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-03-20 12:25 AM					Message:Temperature Alert: 0.97 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-20 12:05 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.94 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-03-20 12:05 AM					Message:Temperature Alert: 0.94 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
2024-03-19 10:09 AM	Sudler 2012 CTSU Drug Fridge	Temperature	Delp G040 +4	1.03 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-19 10:09 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.03 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-03-19 10:09 AM					Message:Temperature Clear: 1.03 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-18 10:42 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.84 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-03-18 10:42 PM					Message:Temperature Alert: 0.84 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-18 10:20 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.97 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-03-18 10:20 PM					Message:Temperature Alert: 0.97 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-18 10:00 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.90 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-03-18 10:00 PM					Message:Temperature Alert: 0.90 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
2024-03-18 9:43 PM	Sudler 2012 CTSU Drug Fridge	Temperature	Delp G040 +4	1.02 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-18 9:43 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.02 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
	2024-03-18 9:43 PM					Message:Temperature Clear: 1.02 °C, within the limit of 1.0 °C to 8.0 °C
						Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-18 9:36 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.00 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
	2024-03-18 9:36 PM					Message:Temperature Alert: 1.00 °C, below the limit of 1.0 °C
						Alert Rule:Delp G040 +4
2024-03-18 1:13 PM	Sudler 2012 CTSU Drug Fridge	Temperature	Delp G040 +4	1.07 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-18 1:13 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.07 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-03-18 1:13 PM					Message:Temperature Clear: 1.07 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-17 9:22 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.87 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-03-17 9:22 AM					Message:Temperature Alert: 0.87 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-17 9:01 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.95 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-03-17 9:01 AM					Message:Temperature Alert: 0.95 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-17 8:38 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.00 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-03-17 8:38 AM					Message:Temperature Alert: 1.00 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4

2024-03-17 8:20 AM	Sudler 2012 CTSU Drug Fridge	Temperature	Delp G040 +4	1.07 °C	8.0/1.0 °C	Alert Cleared
--------------------	------------------------------	-------------	---------------------	---------	------------	---------------

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-17 8:20 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.07 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-03-17 8:20 AM					Message:Temperature Clear: 1.07 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-17 7:53 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.00 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-03-17 7:53 AM					Message:Temperature Alert: 1.00 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
2024-03-13 11:15 AM	Sudler 2012 CTSU Drug Fridge	Temperature	Delp G040 +4	1.55 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-03-13 11:15 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.55 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-03-13 11:15 AM					Message:Temperature Clear: 1.55 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-02-15 12:07 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.84 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-02-15 12:07 PM					Message:Temperature Alert: 0.84 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-02-15 11:46 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.97 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-02-15 11:46 AM					Message:Temperature Alert: 0.97 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-02-15 11:25 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.90 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-02-15 11:25 AM					Message:Temperature Alert: 0.90 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4

CONNECTION ALERTS

No Alert Events for selected date range.

POWER ALERTS

No Alert Events for selected date range.

ERROR ALERTS

No Alert Events for selected date range.

Sudler 2012 CTSU Drug Fridge
for
Sudler 2012 CTSU Drug Fridge

Apr 01, 2024 - May 01, 2024

Author: Miranda Wright
mwright15@kumc.edu

Review & approval records location:

Summary Report "Sudler 2012 CTSU Drug Fridge" for Sudler 2012 CTSU Drug Fridge

Apr 01, 2024 - May 01, 2024


MACHINE DETAILS

Serial Number: **A434F1FF2A93**
Location: **SUDLER 2012 CTSU OUTSIDE LAB**
Description: **A434F1FF2A93**
Installation date:

EQUIPMENT DETAILS

Category: **+4 Refrigerator**
Manufacturer:
Model:
Service Entered Date:

TEMPERATURE AVG: **1.36 °C** | MIN: **0.2 °C** | MAX: **3.71 °C**



DATA LOGS

CENTRAL TIME (US & CANADA)	TEMPERATURE (°C)
2024-03-31 5:00 AM	2.12°C
2024-04-01 5:00 AM	1.88°C
2024-04-02 5:00 AM	1.7°C
2024-04-03 5:00 AM	0.99°C
2024-04-04 5:00 AM	0.73°C
2024-04-05 5:00 AM	0.79°C
2024-04-06 5:00 AM	0.88°C
2024-04-07 5:00 AM	1.41°C
2024-04-08 5:00 AM	1.47°C
2024-04-09 5:00 AM	1.56°C
2024-04-10 5:00 AM	1.46°C
2024-04-11 5:00 AM	1.53°C
2024-04-12 5:00 AM	0.98°C
2024-04-13 5:00 AM	0.98°C
2024-04-14 5:00 AM	2.47°C
2024-04-15 5:00 AM	2.33°C
2024-04-16 5:00 AM	2.19°C
2024-04-17 5:00 AM	2.14°C
2024-04-18 5:00 AM	2.23°C
2024-04-19 5:00 AM	1.3°C
2024-04-20 5:00 AM	1.05°C
2024-04-21 5:00 AM	0.57°C
2024-04-22 5:00 AM	0.81°C
2024-04-23 5:00 AM	0.98°C
2024-04-24 5:00 AM	0.9°C
2024-04-25 5:00 AM	1.23°C
2024-04-26 5:00 AM	1.28°C
2024-04-27 5:00 AM	1.37°C
2024-04-28 5:00 AM	1.73°C
2024-04-29 5:00 AM	1.46°C
2024-04-30 5:00 AM	1.24°C

THRESHOLD ALERTS

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
2024-04-24 4:53 PM	Sudler 2012 CTSU Drug Fridge	Temperature	Delp G040 +4	1.1 °C	8.0/1.0 °C	Alert Cleared
						<p>Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.10 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.</p> <p>Alert Rule:Delp G040 +4</p>
						<p>Message:Temperature Clear: 1.10 °C, within the limit of 1.0 °C to 8.0 °C</p> <p>Alert Rule:Delp G040 +4</p>
						<p>Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.93 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.</p> <p>Alert Rule:Delp G040 +4</p>
						<p>Message:Temperature Alert: 0.93 °C, below the limit of 1.0 °C</p> <p>Alert Rule:Delp G040 +4</p>
						<p>Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.95 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.</p> <p>Alert Rule:Delp G040 +4</p>
						<p>Message:Temperature Alert: 0.95 °C, below the limit of 1.0 °C</p> <p>Alert Rule:Delp G040 +4</p>
						<p>Notification:Email, SMS, and App sent to Miranda Wright.</p> <p>Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.98 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.</p> <p>Alert Rule:Delp G040 +4</p>
						<p>Message:Temperature Alert: 0.98 °C, below the limit of 1.0 °C</p> <p>Alert Rule:Delp G040 +4</p>
2024-04-24 10:59 AM	Sudler 2012 CTSU Drug Fridge	Temperature	Delp G040 +4	1.08 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-24 10:59 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.08 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-04-24 10:59 AM					Message:Temperature Clear: 1.08 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-24 10:48 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.97 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-04-24 10:48 AM					Message:Temperature Alert: 0.97 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
2024-04-24 10:28 AM	Sudler 2012 CTSU Drug Fridge	Temperature	Delp G040 +4	1.03 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-24 10:28 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.03 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
	2024-04-24 10:28 AM					Message:Temperature Clear: 1.03 °C, within the limit of 1.0 °C to 8.0 °C
						Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-23 7:52 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.98 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
	2024-04-23 7:52 PM					Message:Temperature Alert: 0.98 °C, below the limit of 1.0 °C
						Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-23 7:30 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.97 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
	2024-04-23 7:30 PM					Message:Temperature Alert: 0.97 °C, below the limit of 1.0 °C
						Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-23 7:09 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.00 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
	2024-04-23 7:09 PM					Message:Temperature Alert: 1.00 °C, below the limit of 1.0 °C
						Alert Rule:Delp G040 +4
2024-04-23 6:52 PM	Sudler 2012 CTSU Drug Fridge	Temperature	Delp G040 +4	1.03 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-23 12:02 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.08 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
	2024-04-23 12:02 PM					Message:Temperature Clear: 1.08 °C, within the limit of 1.0 °C to 8.0 °C
						Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-23 11:40 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.00 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
	2024-04-23 11:40 AM					Message:Temperature Alert: 1.00 °C, below the limit of 1.0 °C
						Alert Rule:Delp G040 +4
2024-04-23 8:53 AM	Sudler 2012 CTSU Drug Fridge	Temperature	Delp G040 +4	1.1 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-23 8:53 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.10 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-04-23 8:53 AM					Message:Temperature Clear: 1.10 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-22 7:58 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.98 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-04-22 7:58 PM					Message:Temperature Alert: 0.98 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-22 7:37 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.97 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-04-22 7:37 PM					Message:Temperature Alert: 0.97 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-22 7:15 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.00 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-04-22 7:15 PM					Message:Temperature Alert: 1.00 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
2024-04-22 6:58 PM	Sudler 2012 CTSU Drug Fridge	Temperature	Delp G040 +4	1.04 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
----------------------------	--------------	------	------	--------------	--------------	----------

●	2024-04-22 3:16 PM	Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.08 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
●	2024-04-22 3:16 PM	Message:Temperature Clear: 1.08 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4

●	2024-04-22 3:00 PM	Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.99 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
●	2024-04-22 3:00 PM	Message:Temperature Alert: 0.99 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4

●	2024-04-22 2:38 PM	Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.99 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
●	2024-04-22 2:38 PM	Message:Temperature Alert: 0.99 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4

2024-04-22 1:04 PM	Sudler 2012 CTSU Drug Fridge	Temperature	Delp G040 +4	1.13 °C	8.0/1.0 °C	Alert Cleared
●	2024-04-22 1:04 PM	Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.13 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4				
●	2024-04-22 1:04 PM	Message:Temperature Clear: 1.13 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4				
●	2024-04-22 12:55 PM	Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.99 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4				
●	2024-04-22 12:55 PM	Message:Temperature Alert: 0.99 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4				

2024-04-22 12:35 PM	Sudler 2012 CTSU Drug Fridge	Temperature	Delp G040 +4	1.04 °C	8.0/1.0 °C	Alert Cleared
---------------------	------------------------------	-------------	---------------------	---------	------------	---------------

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-22 12:35 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.04 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
	2024-04-22 12:35 PM					Message:Temperature Clear: 1.04 °C, within the limit of 1.0 °C to 8.0 °C
						Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-20 6:28 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.93 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
	2024-04-20 6:28 AM					Message:Temperature Alert: 0.93 °C, below the limit of 1.0 °C
						Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-20 6:06 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.95 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
	2024-04-20 6:06 AM					Message:Temperature Alert: 0.95 °C, below the limit of 1.0 °C
						Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-20 5:45 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.97 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
	2024-04-20 5:45 AM					Message:Temperature Alert: 0.97 °C, below the limit of 1.0 °C
						Alert Rule:Delp G040 +4
2024-04-20 5:25 AM	Sudler 2012 CTSU Drug Fridge	Temperature	Delp G040 +4	1.1 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.13 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-04-20 2:32 AM					Message:Temperature Clear: 1.13 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.98 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-04-20 2:25 AM					Message:Temperature Alert: 0.98 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4

2024-04-13 6:38 AM Sudler 2012 CTSU Drug Fridge Temperature **Delp G040 +4** 1.03 °C 8.0/1.0 °C Alert Cleared

						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.03 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-04-13 6:38 AM					Message:Temperature Clear: 1.03 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4

						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.98 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-04-13 6:30 AM					Message:Temperature Alert: 0.98 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4

2024-04-13 5:29 AM Sudler 2012 CTSU Drug Fridge Temperature **Delp G040 +4** 1.1 °C 8.0/1.0 °C Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.10 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
						Message:Temperature Clear: 1.10 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.98 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
						Message:Temperature Alert: 0.98 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
2024-04-13 2:49 AM	Sudler 2012 CTSU Drug Fridge	Temperature	Delp G040 +4	1.07 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-13 2:49 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.07 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
	2024-04-13 2:49 AM					Message:Temperature Clear: 1.07 °C, within the limit of 1.0 °C to 8.0 °C
						Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-13 1:44 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.97 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
	2024-04-13 1:44 AM					Message:Temperature Alert: 0.97 °C, below the limit of 1.0 °C
						Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-13 1:22 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.98 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
	2024-04-13 1:22 AM					Message:Temperature Alert: 0.98 °C, below the limit of 1.0 °C
						Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-13 1:01 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.99 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
	2024-04-13 1:01 AM					Message:Temperature Alert: 0.99 °C, below the limit of 1.0 °C
						Alert Rule:Delp G040 +4
2024-04-13 12:43 AM	Sudler 2012 CTSU Drug Fridge	Temperature	Delp G040 +4	1.05 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-13 12:43 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.05 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-04-13 12:43 AM					Message:Temperature Clear: 1.05 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-12 11:17 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.94 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-04-12 11:17 PM					Message:Temperature Alert: 0.94 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-12 10:57 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.00 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-04-12 10:57 PM					Message:Temperature Alert: 1.00 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-12 10:34 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.97 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-04-12 10:34 PM					Message:Temperature Alert: 0.97 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
2024-04-12 10:16 PM	Sudler 2012 CTSU Drug Fridge	Temperature	Delp G040 +4	1.09 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-12 10:16 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.09 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-04-12 10:16 PM					Message:Temperature Clear: 1.09 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-12 9:34 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.98 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-04-12 9:34 PM					Message:Temperature Alert: 0.98 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-12 9:12 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.00 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-04-12 9:12 PM					Message:Temperature Alert: 1.00 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-12 8:49 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.99 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-04-12 8:49 PM					Message:Temperature Alert: 0.99 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
2024-04-12 7:54 PM	Sudler 2012 CTSU Drug Fridge	Temperature	Delp G040 +4	1.04 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.08 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
						Message:Temperature Clear: 1.08 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.00 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
						Message:Temperature Alert: 1.00 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4

2024-04-12 4:34 AM Sudler 2012 CTSU Drug Fridge Temperature **Delp G040 +4** 1.04 °C 8.0/1.0 °C Alert Cleared

						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.04 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
						Message:Temperature Clear: 1.04 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4

2024-04-12 4:27 AM Sudler 2012 CTSU Drug Fridge Temperature **Delp G040 +4** 0.98 °C 8.0/1.0 °C Alert Cleared

						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.98 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
						Message:Temperature Alert: 0.98 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4

2024-04-12 4:06 AM Sudler 2012 CTSU Drug Fridge Temperature **Delp G040 +4** 1.08 °C 8.0/1.0 °C Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.08 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
						Message:Temperature Clear: 1.08 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.89 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
						Message:Temperature Alert: 0.89 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4

2024-04-06 6:12 PM Sudler 2012 CTSU Drug Fridge Temperature **Delp G040 +4** 1.13 °C 8.0/1.0 °C Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-06 6:12 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.13 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-04-06 6:12 PM					Message:Temperature Clear: 1.13 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-06 12:08 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.00 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-04-06 12:08 AM					Message:Temperature Alert: 1.00 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-05 11:45 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.97 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-04-05 11:45 PM					Message:Temperature Alert: 0.97 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-05 11:24 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.97 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-04-05 11:24 PM					Message:Temperature Alert: 0.97 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
2024-04-05 11:08 PM	Sudler 2012 CTSU Drug Fridge	Temperature	Delp G040 +4	1.13 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-05 11:08 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.13 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
	2024-04-05 11:08 PM					Message:Temperature Clear: 1.13 °C, within the limit of 1.0 °C to 8.0 °C
						Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-05 10:45 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.99 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
	2024-04-05 10:45 PM					Message:Temperature Alert: 0.99 °C, below the limit of 1.0 °C
						Alert Rule:Delp G040 +4
2024-04-05 10:17 PM	Sudler 2012 CTSU Drug Fridge	Temperature	Delp G040 +4	1.14 °C	8.0/1.0 °C	Alert Cleared
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-05 10:17 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.14 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
	2024-04-05 10:17 PM					Message:Temperature Clear: 1.14 °C, within the limit of 1.0 °C to 8.0 °C
						Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-05 10:12 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.99 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
	2024-04-05 10:12 PM					Message:Temperature Alert: 0.99 °C, below the limit of 1.0 °C
						Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-05 9:50 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.00 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
	2024-04-05 9:50 PM					Message:Temperature Alert: 1.00 °C, below the limit of 1.0 °C
						Alert Rule:Delp G040 +4
2024-04-05 8:39 PM	Sudler 2012 CTSU Drug Fridge	Temperature	Delp G040 +4	1.07 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.07 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
						Message:Temperature Clear: 1.07 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.97 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
						Message:Temperature Alert: 0.97 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.99 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
						Message:Temperature Alert: 0.99 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright. Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.97 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
						Message:Temperature Alert: 0.97 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4

2024-04-05 7:04 PM	Sudler 2012 CTSU Drug Fridge	Temperature	Delp G040 +4	1.07 °C	8.0/1.0 °C	Alert Cleared
--------------------	------------------------------	-------------	---------------------	---------	------------	---------------

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-05 7:04 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.07 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
	2024-04-05 7:04 PM					Message:Temperature Clear: 1.07 °C, within the limit of 1.0 °C to 8.0 °C
						Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-05 12:07 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.57 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
	2024-04-05 12:07 PM					Message:Temperature Alert: 0.57 °C, below the limit of 1.0 °C
						Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-05 11:47 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.62 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
	2024-04-05 11:47 AM					Message:Temperature Alert: 0.62 °C, below the limit of 1.0 °C
						Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-05 11:26 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.79 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
	2024-04-05 11:26 AM					Message:Temperature Alert: 0.79 °C, below the limit of 1.0 °C
						Alert Rule:Delp G040 +4
2024-04-05 11:09 AM	Sudler 2012 CTSU Drug Fridge	Temperature	Delp G040 +4	1.02 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-05 11:09 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.02 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-04-05 11:09 AM					Message:Temperature Clear: 1.02 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-04 11:26 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.98 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-04-04 11:26 PM					Message:Temperature Alert: 0.98 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-04 11:03 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.88 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-04-04 11:03 PM					Message:Temperature Alert: 0.88 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-04 10:42 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.98 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-04-04 10:42 PM					Message:Temperature Alert: 0.98 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
2024-04-04 10:37 AM	Sudler 2012 CTSU Drug Fridge	Temperature	Delp G040 +4	1.03 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-03 11:55 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.03 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-04-03 11:55 PM					Message:Temperature Clear: 1.03 °C, within the limit of 1.0 °C to 8.0 °C Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-03 11:47 PM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.91 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB. Alert Rule:Delp G040 +4
	2024-04-03 11:47 PM					Message:Temperature Alert: 0.91 °C, below the limit of 1.0 °C Alert Rule:Delp G040 +4
2024-04-03 9:03 AM	Sudler 2012 CTSU Drug Fridge	Temperature	Delp G040 +4	1.1 °C	8.0/1.0 °C	Alert Cleared

CENTRAL TIME (US & CANADA)	MACHINE NAME	TYPE	RULE	LAST READING	LIMITS (H/L)	ACTIVITY
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-03 9:03 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 1.10 °C, within the limit of 1.0 °C to 8.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
						Message:Temperature Clear: 1.10 °C, within the limit of 1.0 °C to 8.0 °C
	2024-04-03 9:03 AM					Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-03 4:04 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.98 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
						Message:Temperature Alert: 0.98 °C, below the limit of 1.0 °C
	2024-04-03 4:04 AM					Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-03 3:44 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.97 °C, still below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
						Message:Temperature Alert: 0.97 °C, below the limit of 1.0 °C
	2024-04-03 3:44 AM					Alert Rule:Delp G040 +4
						Notification:Email, SMS, and App sent to Miranda Wright.
	2024-04-03 3:23 AM					Message:Sudler 2012 CTSU Drug Fridge is reporting Temperature of 0.99 °C, below the limit of 1.0 °C. Sudler 2012 CTSU Drug Fridge is located in SUDLER 2012 CTSU OUTSIDE LAB.
						Alert Rule:Delp G040 +4
						Message:Temperature Alert: 0.99 °C, below the limit of 1.0 °C
	2024-04-03 3:23 AM					Alert Rule:Delp G040 +4

CONNECTION ALERTS

No Alert Events for selected date range.

POWER ALERTS

No Alert Events for selected date range.

ERROR ALERTS

No Alert Events for selected date range.