

AED/Power Vac Performance Report														
Customer: University of KS Health Systems CTSU				Contact: Shelby Williams										
Street: 4350 Shawnee Mission Pkwy				Phone #: 913-574-3018										
City/ST/Zip: Fairway, KS 66205				P= Pass F=Fail										
Asset#	Model#	MFG	S/N	Pad Ex. Date	CPR-D Padz EX-Date	Condition Controls/Switches	Manual Self Test	Hardware Integrity	Indicators & Display Operation	MFG. Labeling Intact	Batteries Replaced Y/N Date Due		Calibration Performed	Calibration Due
034856	197	Zoll	AA128023827	2026-16-04	P	P	P	P	P	P	N	EX-03/2025	1/26/2023	01/2024
034857	197	Zoll	AA128023825	2026-16-04	P	P	P	P	P	P	N	EX-03/2025	1/26/2023	01/2024
Power Vacuum				<25 inHgVac										
033860	PM61	Power vac	002535	Yes		P	P	P	P	P	NA	NA	1/26/2023	1/2024
033844	PM61	Power vac	002538	Yes		P	P	P	P	P	NA	NA	1/26/2023	1/2024

Technicians: Test equipment used:

Initials:

Notes:

Calibration Performed by Cobb Refrigeration and Lab Service 8071 Flint St. Lenexa KS, 66212 913-403-0088