

Service Report



WORK ORDER ID: CWKD8232895

WORK ORDER TYPE: Preventative Maintenance

CUSTOMER: UNIVERSITY OF KS MED CTR-CTSU | **DEPARTMENT:** 146148-UNIVERSITY OF KS MED CTR-CTSU-BIOMED | **ACCOUNT NUMBER:** 146148

CONTACT:

CONTACT PHONE:

BILLING ADDRESS: 4350 Shawnee Mission Parkway Mail Stop 6011 Fairway, KS 66205

CUSTOMER PO#:

CUSTOMER REFERENCE:

SHIPPING ADDRESS: 4350 Shawnee Mission Parkway Mail Stop 6011 Fairway, KS 66205

AGILITI ASSET ID : 1705469 | **SERIAL #:** V1534616 | **VENDOR SITE ID:** | **CUSTOMER ASSET ID:**

ASSET TYPE: STERILIZING UNITS, STEAM, TABLETOP | **MANUFACTURER:** MIDMARK CORP | **MODEL:** M11-040

SERVICE REQUESTED: Asset Onboarding - Initial Preventative Maintenance for 1705469

SERVICE PROVIDED TYPE: Scheduled Maintenance

WORK COMPLETED: 04-28-2025

LABOR

| SERVICE PROVIDER | CATEGORY | HOURS | PERIOD DEFINITION | SERVICE DATE |
|------------------|----------------|-------|-------------------|--------------|
| Stephanie Dye | PM Not Covered | 0.5 | M-F 8-5 | 04/28/25 |

SERVICE PROVIDED DETAILS

UNIT PASSED ALL REQUIRED TESTING AT THIS TIME

CHECKLIST

PROCEDURE: AUTOC L1

QUESTION

ADJUST TEMPERTURE

ANSWER

Yes

| | |
|------------------------------|------|
| CUST. CALLED | Yes |
| Electrical Safety Inspection | Pass |
| GAUGES | Pass |
| Mechanical Safety Inspection | Pass |
| OPERATION | Pass |
| OVERPRESSURE | Pass |
| OVERTEMP | Pass |
| PRESSURE DOOR | Pass |
| PROOF PLUS | Pass |
| SAFETY CHECK | Pass |
| SAFETY SWITCH | Pass |
| TIMER | Pass |

Customer Signature:

Date:

Technician Signature:

Date: