

CTSU Lab Room Freezer Min/Max Temperature Log

Maintain temperature between 2°C to 8°C

Refrigerator- Delp
Isotemp 1150957901190626
Refrigerator or Freezer ID/Location:
CTSU RAINBOW

Month 5 May Year 2025 Threshold: 30 minutes

Document the details of any temperature excursion in comments below. Alert study team if excursion exceeds threshold.

	Date (Weekdays only)	Time	Actual Temp	Min °C	Max °C	Initials	Comments
Thu	5/1/2025	1127	8.0	3.4	9.0	MW	
Fri	5/2/2025	0810	4.5	3.3	9.4	MW	
Mon	5/5/2025	0805	4.5	3.3	9.4	MW	
Tue	5/6/2025	0815	4.5	3.3	9.4	MW	
Wed	5/7/2025	0803	7.5	3.4	9.1	MW	
Thu	5/8/2025	0810	7.9	3.3	9.0	MW	
Fri	5/9/2025	0838	7.9	3.3	9.0	MW	
Mon	5/12/2025	1000	5.4	3.3	9.5	DH	
Tue	5/13/2025	0813	5.5	3.3	9.5	MW	
Wed	5/14/2025	0909	7.3	0.6	9.3	MW	
Thu	5/15/2025	0744	3.7	3.4	9.3	MW	
Fri	5/16/2025	0833	5.2	3.5	9.3	MW	
Mon	5/19/2025	0733	7.1	3.4	9.5	MW	
Tue	5/20/2025	0800	8.4	3.5	9.8	MW	
Wed	5/21/2025	0800	7.7	3.6	8.9	ADK	
Thu	5/22/2025	0830	6.2	3.4	8.9	DH	
Fri	5/23/2025	0800	8.7	3.5	9.3		ERROR ADK HOLIDAY
Mon	5/26/2025		8.7	3.5	9.3	ADK	ERROR HOLIDAY
Tue	5/27/2025	0800	7.1	3.4	8.9	ADK	
Wed	5/28/2025	0920	8.4	3.5	9.3	DH	
Thu	5/29/2025	0815	4.6	3.4	8.9	DH	
Fri	5/30/2025	0745	5.6	3.4	8.9	DH	
Mon	6/2/2025						

Key: * Heavy usage