

COBB Refrigeration & Laboratory Services LLC
8071 Flint St, Lenexa, Kansas 66214

Phone 913 403-0088

Fax 913 403-0089

CERTIFICATION OF ACCURACY

Customer Information

Company Name: KU Med Center CTSU

Company Address: 3901 Rainbow, Kansas City, KS

Contact Person/ Phone Number: Danielle Hoage 913-574-3059

Thermometer Information

Manufacturer: Welch Allyn

Model: SureTemp Plus

Serial No: (21)19147495

Department: CTSU

	High	Medium	Low
Certified Temperature (°C):	41.3	38.8	36.8
Observed Temperature (°C):	41.4	39.3	36.8
Observed Accuracy (°C):	0.1	0.5	0

Passed/Failed: Pass

Certification Date: 1/10/2024

Certification Due: 1/2025

Certifying Equipment

Manufacturer: Fluke

Model: 52II

Serial No: 15100368

Due Date: 2/1/2024

Service Technician: Derek Hulett
(Print)

Date: 1/10/2024

Service Technician:
(Sign)

