

CTSU Lab Room Freezer Min/Max Temperature Log

Maintain temperature between 35°C to 37°C

Incubator - Sudler

Boekel 192423685

Refrigerator or Freezer ID/Location:
CTSU RAINBOW

Month 2 February Year 2025 Threshold: 30 minutes

Document the details of any temperature excursion in comments below. Alert study team if excursion exceeds threshold.

	Date (Weekdays only)	Time	Actual Temp	Min °C	Max °C	Initials	Comments
Mon	2/3/2025	0821	36.7	35.5	37.1	PL	
Tue	2/4/2025	0800	37	35.5	37.4	PL	
Wed	2/5/2025	0740	36.3	35.5	37.4	MW	
Thu	2/6/2025	0738	36.3	35.3	37.4	MW	
Fri	2/7/2025	0738	36.7	34.6	37.4	PL	
Mon	2/10/2025	0758	36.7	34.6	37.4	PL	
	2/11/2025	0749	36.7	34.6	37.4	PL	
Wed	2/12/2025						Closed due to Weather
Thu	2/13/2025	0749	36.7	34.6	37.4	PL	
Fri	2/14/2025	0744	36.2	34.6	37.4	PL	
Mon	2/17/2025	0800	36.4	34.6	37.4	PL	
Tue	2/18/2025	Closed					Weather
Wed	2/19/2025	0800	36.2	34.6	37.4	PL	
Thu	2/20/2025	0800	36.2	34.6	37.4	PL	
Fri	2/21/2025	0800	36.7	34.4	37.4	MW	
Mon	2/24/2025	0805	36.9	34.4	37.2	MW	
Tue	2/25/2025	0947	38.3	36.8	38.4	PL	
Wed	2/26/2025	0819	38.9	36.8	39.0	PL	
Thu	2/27/2025	0737	38.8	36.8	39.1	MW	
Fri	2/28/2025	0715	36.2	35.2	38.8	MW	
Mon	3/3/2025						
Tue	3/4/2025						
Wed	3/5/2025						

Key: * Heavy usage