

CTSU Lab Room Freezer Min/Max Temperature Log

Maintain temperature between 2°C to 8°C

Refrigerator- Delp  
Isotemp 1150957901190626

Refrigerator or Freezer ID/Location:  
CTSU RAINBOW

Month 12 December Year 2025 Threshold: 30 minutes

Document the details of any temperature excursion in comments below. Alert study team if excursion exceeds threshold.

Date (Weekdays only)	Time	Actual Temp	Min °C	Max °C	Initials	Comments
Mon 12/1/2025	0835	6.5	-0.5	8.9	MW	
Tue 12/2/2025	0814	7.3	-0.7	8.3	MW	
Wed 12/3/2025	0815	5.9	-0.6	8.5	ADK	
Thu 12/4/2025	0740	4.4	-0.6	8.5	m	
Fri 12/5/2025	0716	0.6	-0.8	8.2	m	
Mon 12/8/2025	0815	6.9	-0.8	8.8	ADK	
Tue 12/9/2025	0748	1.8	-0.8	8.8	m	
Wed 12/10/2025	0839	6.0	-0.6	8.9	w	
Thu 12/11/2025	0800	3.6	-0.5	8.7	ADK	
Fri 12/12/2025	0820	4.4	-0.6	8.7	ADK	
Mon 12/15/2025	0757	2.3	-0.4	8.7	m	
Tue 12/16/2025	0815	7.2	-0.6	8.4	ADK	
Wed 12/17/2025	0805	7.8	-0.6	8.7	m	
Thu 12/18/2025	1520	7.7	-1.5	8.5	w	
Fri 12/19/2025	1515	7.7	-1.5	8.5	w	
Mon 12/22/2025	0801	4.2	1.6	8.6	w	
Tue 12/23/2025	0809	4.2	1.4	8.6	w	
Wed 12/24/2025	0802	5.5	1.4	8.6	w	
Thu 12/25/2025	N/A	N/A	N/A	N/A	m	Closed for the holiday
Fri 12/26/2025						
Mon 12/29/2025						
Tue 12/30/2025						
Wed 12/31/2025						

Key: \* Heavy usage