

CTSU Lab Room Freezer Min/Max Temperature Log

Maintain temperature between -20°C to -30°C

-20 Accucold Freezer - Delp

2021 12 000719

Refrigerator or Freezer ID/Location:
CTSU RAINBOW

Month 4 April Year 2025 Threshold: 4 hours

Document the details of any temperature excursion in comments below. Alert study team if excursion exceeds threshold.

	Date (Weekdays only)	Time	Actual Temp	Min °C	Max °C	Initials	Comments
Tue	4/1/2025	0739	-26.2	-29.2	-22.5	PL	
Wed	4/2/2025	0754	-28.6	-29.4	-22.5	PL	
Thu	4/3/2025	0737	-22.9	-29.4	-22.5	MW	
Fri	4/4/2025	0747	-24.9	-28.5	-17.7	MW	
Mon	4/7/2025	0747	-28.5	-28.5	-17.7	MW	
Tue	4/8/2025	0730	-24.2	-26.6	-23.6	PL	
Wed	4/9/2025	0745	-24.7	-26.8	-23.6	PL	
Thu	4/10/2025	0742	-25.2	-26.9	-23.6	MW	
Fri	4/11/2025	0735	-25.3	-26.8	-24.4	MW	
Mon	4/14/2025	0749	-24.7	-27.2	-24.2	PL	
Tue	4/15/2025	0749	-25	-27.2	-24.2	PL	
Wed	4/16/2025	0835	-24	-27.2	-24.2	ADK	
Thu	4/17/2025	0826	-25.2	-26.9	-24.5	ADK	
Fri	4/18/2025	0750	-26	-27.1	-22.9	MW	
Mon	4/21/2025	0749	-25.2	-26.9	-23.9	PL	
Tue	4/22/2025	0749	-25	-26.9	-23.9	PL	
Wed	4/23/2025	0715	-24.4	-26.9	-20.3	MW	
Thu	4/24/2025	0720	-24.4	-26.9	-20.3	MW	
Fri	4/25/2025	0737	-24.4	-26.9	-20.3	MW	
Mon	4/28/2025	0853	-25.3	-26.9	-23.9	MW	
Tue	4/29/2025	0810	-24.8	-26.9	-23.9	MW	
Wed	4/30/2025	0830	-24.8	-26.9	-23.9	MW	
Thu	5/1/2025						

Key: * Heavy usage