8071 Flint St, Lenexa, Kansas 66214

Phone 913 403-0088 Fax 913 403-0089

CERTIFICATION OF ACCURACY

Customer Information

Company Name: University of KS Medical Center – CTSU Rainbow

Company Address: 3901 Rainbow Blvd Kansas City, KS 66160

Contact Person/ Phone Number: Nelda Mallett 913-574-3007

Thermometer Information: Built-In Manufacturer: Coviden

Model: 10011658 Serial No: A1952353X

Certified Temperature (°C): 40.4

Observed Temperature (°C): 41.1

Observed Accuracy (°C): 0.7

Passed/Failed: Pass

Certification Date: 1/7/2025 Certification Due: 1/2026

Certifying Equipment

Manufacturer: Fluke Model: 52II

Serial No: 15100368 Due Date: 2/2/2025

Service Technician: Derek Hulett Date: 1/7/2025

(Print)

Service Technician:

8071 Flint St, Lenexa, Kansas 66214

Phone 913 403-0088 Fax 913 403-0089

CERTIFICATION OF ACCURACY

Customer Information

Company Name: University of KS Medical Center – CTSU Rainbow

Company Address: 3901 Rainbow Blvd Kansas City, KS 66160

Contact Person/ Phone Number: Nelda Mallett 913-574-3007

Thermometer Information: Built-In Manufacturer: Coviden

Model: 10011658 Serial No: A1952371X

Certified Temperature (°C): 40.5

Observed Temperature (°C): 40.8

Observed Accuracy (°C): 0.3

Passed/Failed: Pass

Certification Date: 1/7/2025 Certification Due: 1/2026

Certifying Equipment

Manufacturer: Fluke Model: 52II

Serial No: 15100368 Due Date: 2/2/2025

Service Technician: Derek Hulett Date: 1/7/2025

(Print)

Service Technician:

8071 Flint St, Lenexa, Kansas 66214

Phone 913 403-0088 Fax 913 403-0089

CERTIFICATION OF ACCURACY

Customer Information

Company Name: University of KS Medical Center – CTSU Rainbow

Company Address: 3901 Rainbow Blvd Kansas City, KS 66160

Contact Person/ Phone Number: Nelda Mallett 913-574-3007

Thermometer Information: Built-In Manufacturer: WelchAllyn

Model: 901053 Serial No: (21)19147463

Certified Temperature (°C): 40.2

Observed Temperature (°C): 39.4

Observed Accuracy (°C): 0.8

Passed/Failed: Pass

Certification Date: 1/7/2025 Certification Due: 1/2026

Certifying Equipment

Manufacturer: Fluke Model: 52II

Serial No: 15100368 Due Date: 2/2/2025

Service Technician: Derek Hulett Date: 1/7/2025

(Print)

Service Technician:

8071 Flint St, Lenexa, Kansas 66214

Phone 913 403-0088 Fax 913 403-0089

CERTIFICATION OF ACCURACY

Customer Information

Company Name: University of KS Medical Center – CTSU Rainbow

Company Address: 3901 Rainbow Blvd Kansas City, KS 66160

Contact Person/ Phone Number: Nelda Mallett 913-574-3007

Thermometer Information: Built-In Manufacturer: WelchAllyn

Model: 901053 Serial No: (21)19147495

Certified Temperature (°C): 40.1

Observed Temperature (°C): 40.0

Observed Accuracy (°C): 0.1

Passed/Failed: Pass

Certification Date: 1/7/2025 Certification Due: 1/2026

Certifying Equipment

Manufacturer: Fluke Model: 52II

Serial No: 15100368 Due Date: 2/2/2025

Service Technician: Derek Hulett Date: 1/7/2025

(Print)

Service Technician: