

Service Report



WORK ORDER ID: CWKD7936859

WORK ORDER TYPE: Preventative Maintenance

CUSTOMER: UNIVERSITY OF KS MED CTR-CTSU | **DEPARTMENT:** 146148-UNIVERSITY OF KS MED CTR-CTSU-BIOMED | **ACCOUNT NUMBER:** 146148

CONTACT:

CUSTOMER PO#:

CONTACT PHONE:

CUSTOMER REFERENCE:

BILLING ADDRESS: 4350 Shawnee Mission Parkway Mail Stop 6011 Fairway, KS 66205

SHIPPING ADDRESS: 4350 Shawnee Mission Parkway Mail Stop 6011 Fairway, KS 66205

AGILITI ASSET ID : 1887921 | **SERIAL #:** 100093420821 | **VENDOR SITE ID:** | **CUSTOMER ASSET ID:**

ASSET TYPE: MONITORS, PHYSIOLOGIC, VITAL SIGNS | **MANUFACTURER:** WELCH ALLYN INC | **MODEL:** CONNEX SPOT MONITOR

SERVICE REQUESTED: PM DUE

SERVICE PROVIDED TYPE: Scheduled Maintenance

WORK COMPLETED: 02-25-2025

LABOR

SERVICE PROVIDER	CATEGORY	HOURS	PERIOD DEFINITION	SERVICE DATE
Stephanie Dye	PM Not Covered	0.4	M-F 8-5	04/28/25

SERVICE PROVIDED DETAILS

UNIT PASSED ALL REQUIRED TESTING AT THIS TIME

TEST EQUIPMENT

PREFIX	UNIT	DESCRIPTION	CALIBRATION DATE
SP4	0105	SPO2 SIMULATOR, FLUKE SPOTLIGHT,PROSIM 4	04/03/25
SP3	0105	VITAL SIGNS PATIENT SIMULATOR, PROSIM 3	04/03/25

CHECKLIST

PROCEDURE: MONITORS, BEDSIDE, PHYSIOLOGIC

QUESTION

ANSWER

01. Check AC plug/line cord/strain relief	Pass
02. Measure chassis ground resistance (milliOhms)	42
03. Measure chassis leakage current (microAmps)	60
04. Patient Leakage (microAmps)	N/A
05. Check outercase, line cords, rolling stands, wall mounts, modular accessories and interconnecting cables	Pass
06. Check patient interface connections (ECG, IBP, SPO2, Temp, CO2 and NIBP)	Pass
07. Check display panel, feet, screws and handle	Pass
08. Keypad/Control Knob Test	Pass
09. Recorder Test	Pass
10. Display Tests	Pass
11. Pixel Test	Pass
12. Color Test	Pass
13. NIBP Tests - Overpressure Voltage Test	Pass
14. NIBP Tests - Static Pressure Calibration	Pass
15. NIBP Tests - Motor Pump Test	Pass
16. NIBP Tests - Leak Test	Pass
17. NIBP Tests - Dump Test (Adult, Neonate)	Pass
18. NIBP Tests - Bleed Rate Test	Pass
19. NIBP Tests - Overpressure Test (Adult, Pediatric, Neonate)	Pass
20. Voltage and Battery Operation Verification	Pass
21. ECG Tests - Initialization	NA
22. ECG Tests - Leads Off	NA
23. ECG Tests - Pacer Detect	NA
24. ECG Tests - Heart Rate	NA
25. ECG Tests - Alarms	NA
26. IBP 1 and IBP 2 Verification	Pass
27. Temperature Verification	Pass
28. SPO2 Verification	Pass
29. SPO2 Alarm	Pass
30. NIBP Verification	Pass
31. CO2 Operation Verification	NA

Customer Signature:

Date:

Stephanie Dye

4/28/25

Technician Signature:

Date: