

COBB Refrigeration & Laboratory Services LLC

8071 Flint St, Lenexa, Kansas 66214

Phone 913 403-0088

Fax 913 403-0089

Centrifuge Performance Report

Customer Information:

Company Name: KU Med Center

Company Address: 3901 Rainbow, Kansas City MO

Contact Person/ Phone Number: Shelby Williams, 913-574-3018, swilliams46@kumc.edu

Equipment Information:

Asset: S-9

Manufacturer: Eppendorf

Model: 5702R

Serial No:

Physical Condition:

Chassis/ Housing:	Pass X	Fail
Controls & Switches:	Pass X	Fail
Power Cord & Strain Relief:	Pass X	Fail
Hardware Integrity:	Pass X	Fail
Indicators & Display Operation:	Pass X	Fail
Mfg. Labeling Intact:	Pass X	Fail
Condenser Filter Replaced:	Pass X	Fail

Operation:

Timer Accuracy	Pass:	Pass 10:00	10:02	Fail:	_____
Lid Latch Safety Interlock	Pass:	Pass	Fail:	_____	_____
Brake	Pass:	Pass	Fail:	_____	_____
Speed Check Low	Set:	300	Measured:	300	_____
Speed Check Med	Set:	2200	Measured:	2193	_____
Speed Check High	Set:	4400	Measured:	4400	_____
Temperature	Set:	4°C	Measured:	5.1°C	_____

Certification Date: 1/11/24

Certification Due: 1/25

Certifying Equipment

Manufacturer: Traceable Stopwatch T-54 Serial:230104133 Due:2/14/25

Manufacturer: ExTech Phototachometer Serial: H437785 Due: 3/23/24

Manufacturer: Fluke 54II Serial: 86470112 Due: 4/18/24

Method: Certified per manufacturer's specifications and procedures traceable to NIST

Service Technician: Nick Brown

Date: 1/11/24

Service Technician: 