

Scale Performance Report

Customer: KU-CTSU Fairway		Contact: Shelby Willams													
Street: 4350 Shawnee Mission Parkway 3rd Floor		Phone # 913-574-3018													
City/ST/Zip: Fairway, KS 66205		P= Pass F=Fail													
MFG	Model#	S/N	Electrical Safety Check	Condition Controls/ Switches	Condition Power Cord & Strin Relief	Hardware Integrity	Indecators & Display Operation	MFG. Labeling Intact	Units of Measure: lbs						Date Performed
									Target	Actual	Target	Actual	Target	Actual	
Health-O-Meter	500KL	5000087608	P	P	P	P	P	P	50.0	50.0	250.0	250.0	500.0	499.8	01/25/2023
Health-O-Meter	500KL	5000135815	P	P	P	P	P	P	50.0	50.0	250.0	249.8	500.0	499.4	01/25/2023
Health-O-Meter	500KL	5000133809	P	P	P	P	P	P	50.0	50.0	250.0	249.6	500.0	499.4	01/25/2023
Detecto	MV1	E14019-0058	P	P	P	P	P	P	50.0	50.0	250.0	250.0	500.0	499.6	01/25/2023

Technicians	JC
Initials	

Technician: Jim Cowart

Test equipment used

50lb Weights: Kansas Metrology Laboratory Report
 Number: 15825 Expires: 02/2023

Notes: Next Due for Calibration 01/2023

