

COBB Refrigeration & Laboratory Services LLC

8071 Flint St, Lenexa, Kansas 66214

Phone 913 403-0088

Fax 913 403-0089

Scale Performance Report

Customer Information

Company Name: University of KS Medical Center- CTSU

Company Address: 3901 Rainbow

Contact Person/ Phone Number: Shelby Williams

Phone Number: 913-574-3018

Manufacturer:

Serial No:

Asset:

Physical Condition:

| | | | |
|---------------------------------|------|----|------|
| Chassis/ Housing: | Pass | X | Fail |
| Controls & Switches: | Pass | Na | Fail |
| Power Cord & Strain Relief: | Pass | Na | Fail |
| Hardware Integrity: | Pass | X | Fail |
| Indicators & Display Operation: | Pass | X | Fail |
| Mfg. Labeling Intact: | Pass | X | Fail |
| Operation: | Pass | X | Fail |

Operation:

| | | |
|-----------------|---------|---------------------|
| Measured Values | Set: 5' | Measured: 4' 31/32" |
| | Set: 4' | Measured: 3' 31/32" |
| | Set: 3' | Measured: 2' 31/32" |

Certification Date: 1/12/2024

Certification Due: 1/12/2025

Certifying Equipment

Method: Certified per manufacturer's specifications using a laser distant measurer

Manufacturer: Bosch Professional GLM 40 S. 506 4000016

Service Technician: Nicholas Brown

Date: 1/12/2024

Service Technician: 
(Sign)