

Scale Performance Report

Customer: KU-CTSU Rainbow		Contact: Shelby Willams													
Street: 3901 Rainbow Blvd (2nd Floor)		Phone #: 913-574-3018													
City/ST/Zip: Kansas City, KS 66160		P= Pass F=Fail													
MFG	Model#	S/N	Electrical Safety Check	Condition Controls/ Switches	Condition Power Cord & Strin Relief	Hardware Integrity	Indecators & Display Operation	MFG. Labeling Intact	Target	Actual	Target	Actual	Target	Actual	Date Performed
Health-O-Meter	500KL	5000163137	P	P	P	P	P	P	50.0	50.0	250.0	250.4	500.0	501.4	01/25/2023
Detecto	MV1	E17619-0049	P	P	P	P	P	P	50.0	50.2	250.0	251.0	500.0	500.6	01/25/2023
Health-O-Meter	500KL	5000178353	P	P	P	P	P	P	50.0	50.0	250.0	250.0	500.0	500.2	01/25/2023

<u>DELP G040</u>															
Detecto	500KL	5000186170	P	P	P	P	P	P	50.0	50.0	250.0	249.6	500.0	499.2	01/25/2023
Health-O-Meter	500KL	5000163141	P	P	P	P	P	P	50.0	50.0	250.0	250.2	500.0	500.0	01/25/2023

Technicians	JC	Technician: Jim Cowart	Test equipment used
Initials			50lb Weights: Kansas Metrology Laboratory Report Number: 15825 Expires: 02/2023

Notes: Next Due for Calibration 01/2024

