

CTSU Lab Room Freezer Min/Max Temperature Log

Maintain temperature between -65°C to -90°C

1124092101210920 (-80 Freezer Sudler)

Refrigerator or Freezer ID/Location:
CTSU RAINBOW

Month 5 May Year 2025 Threshold: 4 hours

Document the details of any temperature excursion in comments below. Alert study team if excursion exceeds threshold.

	Date (Weekdays only)	Time	Actual Temp	Min °C	Max °C	Initials	Comments
Thu	5/1/2025	0715	-81	-79	-81	MW	
Fri	5/2/2025					MW	1030 Defrost Samples moved @ 0900 MW
Mon	5/5/2025	0715	-81	-78	-81	MW	
Tue	5/6/2025	0705	-81	-78	-81	MW	
Wed	5/7/2025	0826	-81	-81	-81	MW	
Thu	5/8/2025	0737	-80	-80	-80	MW	
Fri	5/9/2025	0715	-80	-80	-80	MW	
Mon	5/12/2025	0810	-80	-79	-80	MW	
Tue	5/13/2025	0820	-80	-79	-80	MW	
Wed	5/14/2025	0715	-80	-79	-81	MW	
Thu	5/15/2025	0720	-81	-79	-81	MW	
Fri	5/16/2025	0727	-81	-79	-81	MW	
Mon	5/19/2025	0710	-80	-79	-81	MW	
Tue	5/20/2025	0713	-80	-79	-81	MW	
Wed	5/21/2025	0735	-81	-79	-81	MW	
Thu	5/22/2025	0800	-81	-81	-81	ADK	
Fri	5/23/2025	0800	-79	-79	-81	ADK	
Mon	5/26/2025	0	-81	-79	-81	ADK	ERROR ADK Holiday
Tue	5/27/2025	0800	-81	-79	-81	ADK	
Wed	5/28/2025	1130	-79	-79	-79	ADK	
Thu	5/29/2025	0800	-81	-77	-81	ADK	
Fri	5/30/2025	0800	-81	-77	-81	ADK	
Mon	6/2/2025						

Key: * Heavy usage