

# CERTIFICATION OF ACCURACY



## Customer Information:

Company Name: CTSU Fairway

Company Address: 4350 Shawnee Mission Pkwy, Fairway, KS, 66205

Contact Person: Shelby Williams

Phone: (913) 574-3018

Thermometer Information: Built-in

Manufacturer: Midmark

Model: M9-022

Serial No: V1534616

Department/Location: 3<sup>rd</sup> Floor

Asset#: M9

Temperature Range(°F): 272

Time Range (m:ss): 15:00

Pressure Range (psi): 30

	<u>Temperature (°F):</u>	<u>Time (m:ss):</u>	<u>Pressure (psi):</u>
Certified Value:	272	15:00	30
Observed Value:	272	15:00	30
Observed Accuracy:	0	0	0

Passed/Failed: Passed

Certification Date: 01/07/2026

Certification Due: 01/2027

## Certifying Equipment:

Thermometer Manufacturer: SP Scienceware

Serial No: 613150000

Model: Durac

Due Date: 7/22/2026

Timer Manufacturer: Traceable

Serial No: 250294394

Model: Stopwatch

Due Date: 05/08/2027

Pressure Gauge Manufacturer: Ashcroft

Serial No: E796099

Model: 2507

Due Date: 04/30/2026

Service Technician: Robert McElwee

(Print)

Service Technician: \_\_\_\_\_

(Sign)

Date: 01/07/2026

