

**COBB Refrigeration & Laboratory Services LLC**

8071 Flint St, Lenexa, Kansas 66214

Phone 913 403-0088

Fax 913 403-0089

**CERTIFICATION OF ACCURACY**

**Customer Information**

**Company Name: University of KS Medical Center-CTSU**

**Company Address: 3901 Rainbow**

**Contact Person: Shelby Williams**

**Phone Number: 913-574-3018**

---

**Thermometer Information: Built-In**

**Manufacturer: Aegis Scientific**

**Model: 2-F-4**

**Serial No: 0820872-38728-Y**

**Department/Location: CTSU Rainbow**

**Asset # N/A**

**Set Point: -22°C**

**Certified Temperature: -19.1°C**

**Observed Temperature: -21.5°C**

**Elemental Monitor: -23.85°C**

**Observed Accuracy: 2.4°C**

**Passed/Failed: Pass**

**Certification Date: 1/10/2024**

**Certification Due: 1/10/2025**

- Unit needs to be defrosted. Excessive ice built up

---

**Certifying Equipment**

**Manufacturer: Fluke**

**Model: 54iiB**

**Serial No: 86470112**

**Due Date: 04/18/2024**

---

**Service Technician: Nicholas Brown**  
**(Print)**

**Date: 1/10/2024**

**Service Technician:**   
**(Sign)**