

# CTSU LAB -8o FREEZER Min/Max Temperature Log

*Maintain temperature between -65°C to -90°C*

Month **March**

Year

2024

Regrigerator ID/Location: **SUDLER 2012E**

	Date (Weekdays only)	Time	Temp °C	Min °C	Max °C	Initials	Comments
Fri	3/1/2024	0845	-81	-81	-81	DH	
Mon	3/4/2024	0842	-81	-81	-81	PL	
Tue	3/5/2024	1400	-79	-79	-79	DH	
Wed	3/6/2024	0800	-81	-81	-81	PL	
Thu	3/7/2024	0830	-80	-81	-81	DH	
Fri	3/8/2024	1000	-79	-80	-80	DH	
Mon	3/11/2024	0934	-80	-80	-81	PL	
Tue	3/12/2024	0617	-79	-79	-80	W	
Wed	3/13/2024	0822	-80	-80	-80	PL	
Thu	3/14/2024	0850	-81	-81	-81	PL	
Fri	3/15/2024	0834	-81	-79	-81	W	
Mon	3/18/2024	0838	-80	-80	-80	PL	
Tue	3/19/2024	0744	-81	-79	-81	MW	
Wed	3/20/2024	0816	-81	-81	-81	PL	
Thu	3/21/2024	0734	-80	-80	-80	MW	
Fri	3/22/2024	0844	-81	-79	-81	W	
Mon	3/25/2024	1156	-81	-80	-80	W	
Tue	3/26/2024	0722	-81	-80	-81	W	
Wed	3/27/2024	0739	-81	-81	-81	W	
Thu	3/28/2024	0743	-81	-81	-81	W	
Fri	3/29/2024	0710	-79	-79	-79	W	

# CTSU Lab -8o FREEZER Min/Max Temperature Log

Maintain temperature between -65°C to -90°C

Month February

Year 2024

Refrigerator Name/  
Location/Serial #/ID Swoller 2012E

	Date (Weekdays only)	Time	Temp °C	Min °C	Max °C	Name/Initials	Comments
Thu	2/1/2024	0630	-81	-81	-81	my/mw	
Fri	2/2/2024	1615	-79	-79	-80	JH	
Mon	2/5/2024	0850	-80	-80	-80	PL	
Tue	2/6/2024	1124	-79	-80	-80	PL	
Wed	2/7/2024	0823	-81	-81	-81	my/mw	
Thu	2/8/2024	0721	-80	-80	-79	mw/ly	
	2/9/2024	0837	-80	-81	-81	mw	
Mon	2/12/2024	0850	-81	-80	-80	PL	
Tue	2/13/2024	0810	-81	-79	-81	PL	
Wed	2/14/2024	0730	-80	-80	-81	mw/ly	
Thu	2/15/2024	0745	-78	-78	-79	<del>mw/ly</del>	
Fri	2/16/2024	0802	-81	-81	-81	my/mw	
Mon	2/19/2024	0815	-79	80	80	PLA	
Tue	2/20/2024	0905	-81	-79	-81	PL/PL	
Wed	2/21/2024	0733	-80	-80	-80	mw/ly	
Thu	2/22/2024	0630	-81	-81	-81	mw/ly	
Fri	2/23/2024	0734	-80	-79	-81	mw/ly	
Mon	2/26/2024	0852	-81	-79	-81	mw/ly	
Tue	2/27/2024	0435	-81	-81	-81	mw/ly	
Wed	2/28/2024	0645	-81	-81	-81	mw/ly	
Thu	2/29/2024	0635	-81	-81	-81	mw/ly	

# CTSU Lab -8o FREEZER Min/Max Temperature Log

Maintain temperature between -65°C to -90°C

Month January

Year 2024

Refrigerator Name/  
Location/Serial #/ID

Sudler 2012

	Date (Weekdays only)	Time	Temp °C	Min °C	Max °C	Name/Initials	Comments	
Tue	1/2/2024	1312	-80	-80	-80	W/mw		
Wed	1/3/2024	0750	-80	-80	-80	W/mw		
Thu	1/4/2024	0749	-81	-81	-81	W/mw		
Fri	1/5/2024	0750	-81	-81	-81	W/mw		
Mon	1/8/2024	0732	-79	-79	-79	W/mw		
Tue	1/9/2024	_____						Closed
	1/10/2024	0715	-80	-80	-80	W/mw		
Thu	1/11/2024	0710	-79	-79	-79	W/mw		
Fri	1/12/2024	0700	-79	-79	-79	W/mw		
Mon	1/15/2024	_____						Holiday
Tue	1/16/2024	1015	-79	-79	-79	W/mw		
Wed	1/17/2024	1330	-78	-79	-79	W/ES		
Thu	1/18/2024	0914	-80	-79	-81	W/PL		
Fri	1/19/2024	0755	-81	-81	-81	W/mw		
Mon	1/22/2024	0720	-78	-80	-78	W/mw		
Tue	1/23/2024	0705	-79	-79	-79	W/mw		
Wed	1/24/2024	0857	-80	-79	-80	W/mw		
Thu	1/25/2024	0647	-81	-81	-81	W/mw		
Fri	1/26/2024	0740	-81	-81	-81	W/mw		
Mon	1/29/2024	0754	-80	-80	-80	W/mw		
Tue	1/30/2024	0430	-81	-81	-81	W/mw		
	1/31/2024	1230	-79	-79	-79	W/mw		

# CTSU LAB -8o FREEZER Min/Max Temperature Log

*Maintain temperature between -65°C to -90°C*

Month **April**

Year

2024

Refrigerator ID/Location: SUDLER 2012E

	Date (Weekdays only)	Time	Temp °C	Min °C	Max °C	Initials	Comments
Mon	4/1/2024	0800	-80	-80	-80	DH	
Tue	4/2/2024	0807	-81	-81	-81	W	
Wed	4/3/2024	0849	-81	-81	-81	W	
Thu	4/4/2024	0815	-81	-81	-81	W	
Fri	4/5/2024	0826	-81	-81	-81	W	
Mon	4/8/2024	0816	-80	-80	-80	W	
Tue	4/9/2024	0800	-80	-79	-80	W	
Wed	4/10/2024	0810	-81	-81	-81	W	
Thu	4/11/2024	1006	-81	-81	-81	W	
Fri	4/12/2024	0711	-81	-81	-81	W	
Mon	4/15/2024	0724	-80	-80	-80	W	
Tue	4/16/2024	0933	-81	-81	-81	PL	
Wed	4/17/2024	0754	-81	-81	-81	W	
Thu	4/18/2024	0716	-81	-81	-81	W	
Fri	4/19/2024	0730	-80	-80	-80	W	
Mon	4/22/2024	0730	-80	-80	-80	W	
Tue	4/23/2024	1222	-77	-77	-77	PL	
Wed	4/24/2024	0720	-81	-80	-81	W	
Thu	4/25/2024	0706	-81	-81	-81	W	
Fri	4/26/2024	0735	-80	-81	-81	W	
Mon	4/29/2024	0710	-80	-80	-80	W	
Tue	4/30/2024	0620	-81	-81	-81	W	