

CTSU Lab Room Freezer Min/Max Temperature Log

Maintain temperature between -20°C to -30°C

-20 Accucold Freezer - Delp

2021 12 000719

Refrigerator or Freezer ID/Location:
CTSU RAINBOW

Month 8 August Year 2025 Threshold: 4 hours

Document the details of any temperature excursion in comments below. Alert study team if excursion exceeds threshold.

	Date (Weekdays only)	Time	Actual Temp	Min °C	Max °C	Initials	Comments
Fri	8/1/2025	0730	-26.3	-28.1	-22.4	MW	
Mon	8/4/2025	0807	-25.3	-28.1	-22.4	MW	
Tue	8/5/2025	0744	-22.3	-28.0	-22.2	MW	
Wed	8/6/2025	0805	-25.1	-27.7	-22.0	MW	
Thu	8/7/2025	0750	-25.2	-27.7	-22.1	ADK	
Fri	8/8/2025	0900	-23.5	-27.8	-22.0	ADK	
Mon	8/11/2025	0750	-26.7	-28.3	-22.0	ADK	
Tue	8/12/2025	0755	-23.7	-28.3	-22.0	ADK	
Wed	8/13/2025	1800	-23.0	-28.0	-22.0	ADK	
Thu	8/14/2025	0830	-24.3	-28.0	-22.6	ADK	
Fri	8/15/2025	0850	-23.0	-28.1	-22.5	ADK	
Mon	8/18/2025	0755	-26.8	-28.2	-22.5	MW	
Tue	8/19/2025	0816	-23.0	-28.2	-22.7	MW	
Wed	8/20/2025	1109	-26.8	-28.1	-22.6	MW	
Thu	8/21/2025	0742	-23.0	-28.1	-22.6	MW	
Fri	8/22/2025	0927	-24.5	-28.9	-20.8	MW	
Mon	8/25/2025	0814	-24.0	-29.2	-22.3	MW	
Tue	8/26/2025	0905	-23.6	-29.0	-22.3	MW	
Wed	8/27/2025	0747	-24.7	-28.8	-22.3	MW	
Thu	8/28/2025	0750	-24.0	-28.6	-22.3	ADK	
Fri	8/29/2025	0730	-23.6	-29.0	-22.3	MW	
Mon	9/1/2025					N/A	
Tue	9/2/2025					N/A	

Key: * Heavy usage