

COBB Refrigeration & Laboratory Services LLC
8071 Flint St, Lenexa, Kansas 66214

Phone 913 403-0088

Fax 913 403-0089

CERTIFICATION OF ACCURACY

Customer Information

Company Name: University of KS Medical Center-CTSU

Company Address: 3901 Rainbow Blvd Kansas City, KS 66103

Contact Person: Shelby Williams

Phone Number: 913-574-3018

Thermometer Information: Built-In

Manufacturer: Thermo ISOTEMP

Model: 05LREEFSA

Serial No: 1150957901190626

Department/Location: CTSU Rainbow

Asset # N/A

Set Point: N/A

Certified Temperature: 4°C

Observed Temperature: N/A

Elemental Monitor: 4.31°C

Observed Accuracy: .31°C

Passed/Failed: Pass

Certification Date: 1/10/2024

Certification Due: 1/10/2025

Certifying Equipment

Manufacturer: Fluke

Model: 54iiB

Serial No: 86470112

Due Date: 04/18/2024

Service Technician: Nicholas Brown

Date: 1/10/2024

(Print)

Service Technician: 

(Sign)