

CTSU Lab Room Freezer Min/Max Temperature Log

Maintain temperature between 35°C to 37°C

Incubator - Sudler
Boekel 192423685

Refrigerator or Freezer ID/Location:
CTSU RAINBOW

Month **3** March Year **2025** Threshold: 30 minutes

Document the details of any temperature excursion in comments below. Alert study team if excursion exceeds threshold.

	Date (Weekdays only)	Time	Actual Temp	Min °C	Max °C	Initials	Comments
Mon	3/3/2025	0115	36.2	35.2	38.8	MW	
Tue	3/4/2025	0752	38.1	35.2	38.8	PL	
Wed	3/5/2025	1212	37.9	35.2	38.8	MW	
Thu	3/6/2025	0737	37	36.6	37.9	MW	
Fri	3/7/2025	0730	37.1	36.5	37.9	MW	
Mon	3/10/2025	0921	37.6	35.5	37.1	PL	
Tue	3/11/2025	0917	37.8	35.5	37.9	PL	
Wed	3/12/2025	0811	37.8	35.5	38.2	MW	
Thu	3/13/2025	0800	37	36	38.4	MW	
Fri	3/14/2025	0800	37	36	38.4	MW	
Mon	3/17/2025	1004	37.5	36	38.4	PL	
Tue	3/18/2025	1000	37.5	36	38.4	PL	
Wed	3/19/2025	1000	37.5	36	38.4	PL	
Thu	3/20/2025	0923	37.7	36	38.4	PL	
Fri	3/21/2025	0700	37	36	38	MW	
Mon	3/24/2025	0800	37.4	36.3	37.6	TH	
Tue	3/25/2025	0814	37.5	37.1	37.6	MW	
Wed	3/26/2025	0755	37.4	37.3	37.8	PL	
Thu	3/27/2025	0815	37.5	37.3	37.8	TH	
Fri	3/28/2025	0900	38.0	37.3	38.2	TH	
Mon	3/31/2025	0840	37.3	37.2	38.5	TH	
Tue	4/1/2025						
Wed	4/2/2025						

Key: * Heavy usage