

Service Report



WORK ORDER ID: CWKD4101265

WORK ORDER TYPE: Preventative Maintenance

CUSTOMER: UNIV OF KS MED CTR-CTSU RNBW | **DEPARTMENT:** 177820-UNIV OF KS MED CTR-CTSU RNBW-ROUNDS CTSU RNBW | **ACCOUNT NUMBER:** 177820

CONTACT:

CUSTOMER PO#:

CONTACT PHONE:

CUSTOMER REFERENCE:

BILLING ADDRESS: 3901 Rainbow Blvd MS 1077 2012 Suder Kansas City, KS 66160 | **SHIPPING ADDRESS:** 3901 Rainbow Blvd MS 1077 2012 Suder Kansas City, KS 66160

AGILITI ASSET ID : 1887947 | **SERIAL #:** 9303139 | **VENDOR SITE ID :** | **CUSTOMER ASSET ID:**

ASSET TYPE: INCUBATORS, LABORATORY, AEROBIC | **MANUFACTURER:** QUINCY LAB | **MODEL:** 10-140

SERVICE REQUESTED: Asset Onboarding - Initial Preventative Maintenance for 18877947

SERVICE PROVIDED TYPE: Scheduled Maintenance

WORK COMPLETED: 01-30-2024

LABOR

SERVICE PROVIDER	CATEGORY	HOURS	PERIOD DEFINITION	SERVICE DATE
Marcus Boatwright	PM Not Covered	0.3	M-F 8-5	01/30/24

SERVICE PROVIDED DETAILS

The door is cracked at the bottom, and the stopper at the top of the device is also breaking. I let the customer know to be careful but it did pass inspection testing 190/407 ES

TEST EQUIPMENT

PREFIX	UNIT	DESCRIPTION	CALIBRATION DATE
SB6	223	SAFETY ANALYZER, FLUKE ESA612	04/25/23
DM1	286	DIGITAL MULTIMETER	04/14/23

CHECKLIST

PROCEDURE: L1200 L1

QUESTION

ANSWER

Alarm - High Temperature

NA

Alarm - Low Temperature

NA

CES THERM.

Pass

Clinical- Audio

Pass

Clinical- Visual

Pass

CUST. THERM.

Pass

DESIRED TEMP

Pass

Electrical Safety Inspection

Pass

Mechanical Safety Inspection

Pass

Operational Check

Pass

Patient- Audio

Pass

Patient- Visual

Pass

Remote

Pass

Safety- Audio

Pass

Safety- Visual

Pass

Customer Signature:

Date:

Technician Signature:

Date: