

CTSU LAB REFRIGERATOR Min/Max Temperature Log

Maintain temperature between 2°C to 8°C

Month **December**

Year **2023**

Refrigerator ID/Location: **SUDLER 2012 LAB**

	Date (Weekdays only)	Time	Temp °C	Min °C	Max °C	Initials	Comments
Fri	12/1/2023	0740	5.7	5.4	5.9	mw	
Mon	12/4/2023	0750	5.6	5.5	5.8	w	
Tue	12/5/2023	0737	5.5	5.4	5.9	mw	
Wed	12/6/2023	0737	5.5	5.4	5.9	mw	
Thu	12/7/2023	0747	5.6	5.4	5.9	mw	
Fri	12/8/2023	0730	5.8	5.5	6.4	mw	
M.	12/11/2023	0826	6	5.8	6.6	w	
Tue	12/12/2023	0824	5.8	5.8	6.1	w	
Wed	12/13/2023	0755	6.0	5.8	6.0	mw	
Thu	12/14/2023		4.3	4.2	4.8		
Fri	12/15/2023		4.3	4.2	4.8		
Mon	12/18/2023	0754	4.3	4.2	6.2	hw	
Tue	12/19/2023	0746	4.4	4.2	4.8	mw	
Wed	12/20/2023	0750	4.5	4	4.9	w	
Thu	12/21/2023	0642	4.4	4	5	w	
Fri	12/22/2023	0705	4.4	4	5.1	w	
Mon	12/25/2023	<hr/>					Closed for the Holiday Break
Tue	12/26/2023	<hr/>					
Wed	12/27/2023	<hr/>					
Thu	12/28/2023	<hr/>					
Fri	12/29/2023	<hr/>					

CTSU Lab Refrigerator Min/Max Temperature Log

Maintain temperature between 2°C to 8 °C

Month **November**

Year **2023**

Refrigerator ID/Location: SUDLER LAB 2012E

Date (Weekdays only)	Time	Temp °C	Min °C	Max °C	Initials	Comments
Wed 11/1/2023	0732	5.3	3.4	5.8	w	
Thu 11/2/2023	0917	5.2	5	5.5	w	
Fri 11/3/2023	0718	5.4	5	5.6	w	
Mon 11/6/2023	0810	5.6	5.4	5.7	PL	
Tue 11/7/2023	0751	5.6	5.3	6	w	
Wed 11/8/2023	0900	4.7	3.8	6.8	DH	
Thu 11/9/2023	1630	4.6	3.8	6.8	DH	
Fri 11/10/2023	—————→					Closed
Mon 11/13/2023	0701	5.8	3.8	6.8	w	
Tue 11/14/2023	0752	5.7	5.6	6	w	
Wed 11/15/2023	1710	4.4	5.6	5.9	DH	
Thu 11/16/2023	0747	5.9	5.6	6.5	w	
Fri 11/17/2023	0721	5.9	5.6	6.5	w	
Mon 11/20/2023	1100	4.4	5.6	6.5	w	
Tue 11/21/2023	1000	4.6	5.9	6.4	DH	
Wed 11/22/2023	0905	6.1	6	6.4	w	
Thu 11/23/2023	—————					Closed
Fri 11/24/2023	—————					Closed
Mon 11/27/2023	0803	5.6	5.4	6.2	w	
Tue 11/28/2023	0735	5.5	5.3	5.7	MW	
Wed 11/29/2023	0731	5.5	5.3	5.7	MW	
Thu 11/30/2023	0730	5.6	5.4	5.9	w	

CTSU Lab Refrigerator Min/Max Temperature Log

Maintain temperature between 2°C to 8 °C

Month **October**

Year

Year **2023**

Refrigerator ID/Location: SUDLER 2012E LAB

	Date (Weekdays only)	Time	Temp °C	Min °C	Max °C	Initials	Comments
Mon	10/2/2023	0738	4	3.1	4.8	W	
Tue	10/3/2023	0948	3.7	3.5	4.6	K	
Wed	10/4/2023	0935	4.3	3.2	4.7	PL	
Thu	10/5/2023	0742	4.1	3.2	4.7	W	
Fri	10/6/2023	0743	4.1	3.2	4.7	W	
Mon	10/9/2023	0811	4.1	3.4	4.7	PL	
	10/10/2023	0814	3.9	3.4	4.8	PL	
Wed	10/11/2023	0840	3.8	3.0	4.8	PL	
Thu	10/12/2023	0833	3.6	2.9	4.8	W	
Fri	10/13/2023	0831	5.2 3.5	2.9	4.5	JUAN	
Mon	10/16/2023	0743	4	2.9	4.5	W	
Tue	10/17/2023	0959	4.3	3.2	4.5	W	
Wed	10/18/2023	0853	3.7	3.5	4.6	W	
Thu	10/19/2023	0948	4.3	3.1	4.6	W	
Fri	10/20/2023	0742	3.9	3.2	4.8	W	
Mon	10/23/2023	0744	3.9	3.2	4.8	W	
Tue	10/24/2023	1013	3.9	3.3	4.7	PL	
Wed	10/25/2023	0716	4.1	3.3	4.7	AW	
Thu	10/26/2023	0825	4.3	2.4	5.3	W	
Fri	10/27/2023	0825	4.3	2.4	5.3	W	
Mon	10/30/2023	0825	4.3	2.4	5.3	W	
Tue	10/31/2023	0821	4.6	3.7	4.6	PL	

CTSU Lab REFRIGERATOR Min/Max Temperature Log

Maintain temperature between 2°C to 8°C

Month **September**

Year **2023**

Refrigerator ID/Location: SUDLER 2012F LAB

Date (Weekdays only)	Time	Temp	Min °C	Max °C	Initials	Comments
Fri 9/1/2023	0752	4	3.4	4.4	w	
Mon 9/4/2023						Closed
Tue 9/5/2023	0812	3.7	3.3	4.5	PL	
Wed 9/6/2023	0848	4.1	4.1	4.1	w	
Thu 9/7/2023	0722	4.3	3.7	4.7	w	
Fri 9/8/2023	0826	4.5	3.6	4.8	PL	
Mon 9/11/2023	0810	4.4	3.6	4.8	mw	
Tue 9/12/2023	0801	4	3.3	4.6	w	
Wed 9/13/2023	0731	4.1	3	5	w	
Thu 9/14/2023	0742	4.1	3.1	5	w	
Fri 9/15/2023	0730	4.4	3.5	4.7	w	
Mon 9/18/2023	0738	3.8	3.5	4.9	w	
Tue 9/19/2023	1340	3.8	3.4	4.7	PL	
Wed 9/20/2023	0730	4.3	3.4	4.7	w	
Thu 9/21/2023	0858	3.8	3.5	4.8	PL	
Fri 9/22/2023	0737	3.8	3.5	4.8	w	
Mon 9/25/2023	0900	4.3	3.3	4.7	PL	
Tue 9/26/2023	1157	4.7	3.3	4.7	DA	
Wed 9/27/2023	1144	4.4	3.3	4.7	PL	
Thu 9/28/2023	0946	5.0	3.2	4.8	DA	
Fri 9/29/2023						

CTSU Lab Refrigerator Min/Max Temperature Log

Maintain temperature between 2°C to 8 °C

Month **August**

Year **2023**

Refrigerator ID/Location: SUDLER 2012E LAB

	Date (Weekdays only)	Time	Temp	Min °C	Max °C	Initials	Comments
Tue	8/1/2023	0607	4.2	3.4	4.3	m	
Wed	8/2/2023	1211	3.9	3.2	4.3	w	
Thu	8/3/2023	0816	3.7	3.2	4.3	m	
Fri	8/4/2023	0800	4.1	3.4	4.3	w	
Mon	8/7/2023	0750	3.7	3.3	5.4	m	
Tue	8/8/2023	0830	4.2	3.2	4.4	PL	
Wed	8/9/2023	0820	3.8	3.2	4.4	w	
	8/10/2023	0832	3.6	3.3	4.3	w	
Fri	8/11/2023	1427	3.4	3.1	4.3	w	
Mon	8/14/2023	0728	3.4	3.2	4.1	w	
Tue	8/15/2023	0628	3.4	3.2	4.1	w	
Wed	8/16/2023	0756	3.7	3	4.1	m	
Thu	8/17/2023	0746	3.5	3.4	4.2	w	
Fri	8/18/2023	0746	3.5	3.4	4.2	w	
Mon	8/21/2023	0747	3.4	3	4.1	w	
Tue	8/22/2023	0956	3.7	3.1	4.4	PL	
Wed	8/23/2023	0822	4.2	3.1	4.5	PL	
Thu	8/24/2023	0822	4.2	3.1	4.5	PL	
Fri	8/25/2023	10:55	3.4	3.1	4.6	PL	
Mon	8/28/2023	0653	3.8	3.1	4.9	w	
Tue	8/29/2023	0743	4.2	3.4	4.4	m	
Wed	8/30/2023	0734	4	3.4	4.8	w	
Thu	8/31/2023	0752	4	3.4	4.4	w	

CTSU Lab Refrigerator Min/Max Temperature Log

Maintain temperature between 2°C to 8 °C

Month **July**

Year **2023**

Refrigerator ID/Location: **SUDLER 2012 LAB**

	Date (Weekdays only)	Time	Temp	Min °C	Max °C	Initials	Comments
Mon	7/3/2023	0816	3.7	3.3	4.5	m	
Tue	7/4/2023	0848	3.9	3.0	4.5	PL	
Wed	7/5/2023	0916	3.9	3.0	4.5	PL	
Thu	7/6/2023	0906	4.9	3.0	4.5	DH	
Fri	7/7/2023	1720	4.9	3.0	5.0	DH	
Mon	7/10/2023	0753	3.9	3.0	5.0	mw	
	7/11/2023	0735	3.8	3.6	4.4	w	
Wed	7/12/2023	0731	3.9	3.6	4.4	w	
Thu	7/13/2023	0746	3.9	3.6	4.4	w	
Fri	7/14/2023	0729	3.9	3.4	4.6	w	
Mon	7/17/2023	0846	3.9	3.3	4.4	w	
Tue	7/18/2023	0854	4.2	3.3	4.4	PL	
Wed	7/19/2023	0854	4.2	3.3	4.4	PL	
Thu	7/20/2023	0715	3.8	3.3	5.2	w	
Fri	7/21/2023	0710	3.8	3.3	5.2	w	
Mon	7/24/2023	0713	3.8	3.3	5.2	w	
Tue	7/25/2023	0815	4.1	3.5	4.1	m	
Wed	7/26/2023	0800	4.2	3.1	4.3	m	
Thu	7/27/2023	0810	4.2	3.1	4.3	w	
Fri	7/28/2023	1354	4.1	3.2	4.3	mw	
Mon	7/31/2023	0854	3.9	3.7	4.3	mw	

CTSU Lab Refrigerator Min/Max Temperature Log

Maintain temperature between 2°C to 8 °C

Month **June**

Year **2023**

Refrigerator ID/Location:

	Date (Weekdays only)	Time	Min °C	Max °C	Initials	Comments
Thu	6/1/2023	4.5	4.4	5	m	
Fri	6/2/2023	4.4	4.1	5	w	
Mon	6/5/2023	4.7	4.4	4.8	w	
Tue	6/6/2023	4.4	4	5.1	w	
Wed	6/7/2023	4.1	2.9	4.8	w	
Thu	6/8/2023	4.1	2.9	4.8	w	
	6/9/2023	3.9	3.3	4.2	m	
Mon	6/12/2023	3.8	3.4	4.3	PL	
Tue	6/13/2023	3.6	2.8	4.3	PL	
Wed	6/14/2023	3.8	2.8	4.4	w	
Thu	6/15/2023	3.7	3.3	4.3	w	
Fri	6/16/2023	3.7	2.8	4.3	w	
Mon	6/19/2023	3.8	3.1	4.3	K	
Tue	6/20/2023	4.1	2.8	4.3	w	
Wed	6/21/2023	3.3	2.5	4.5	m	
Thu	6/22/2023	3.3	2.7	4.3	w	
Fri	6/23/2023	3.3	2.7	4.3	w	
Mon	6/26/2023	4.2	3.2	4.3	w	
Tue	6/27/2023	3.7	2.7	4.5	R	
Wed	6/28/2023	4.2	2.7	4.5	R	
Thu	6/29/2023	3.8	2.7	4.5	PL	
	6/30/2023	3.7	2.7	4.7	w	

CTSU Lab Refrigerator Min/Max Temperature Log

Maintain temperature between 2°C to 8 °C

Month **May**

Year **2023**

Refrigerator ID/Location: *Judler 2012*

	Date (Weekdays only)	Time	Min °C	Max °C	Initials	Comments
Mon	5/1/2023	4.4	3.8	4.8	w	
Tue	5/2/2023	4.2	3.6	4.8	w	
Wed	5/3/2023	4	3.5	4.4	mw	
Thu	5/4/2023	4	3.6	4.4	mw	
Fri	5/5/2023	4	3.6	4.4	mw	
Mon	5/8/2023	5	3.5	5.2	JH	
Tue	5/9/2023	4.4	3.5	5.2	w	
	5/10/2023	4.4	3.5	5.2	w	
Thu	5/11/2023	4.3	3.8	4.6	PL	
Fri	5/12/2023	4.1	3.8	4.7	w	
Mon	5/15/2023	4.1	3.8	4.7	w	
Tue	5/16/2023	4.1	3.9	4.6	w	
Wed	5/17/2023	4.1	3.8	4.6	PL	
Thu	5/18/2023	4.5	3.8	4.9	w	
Fri	5/19/2023	4.1	3.9	4.6	w	
Mon	5/22/2023	4.7	3.8	4.9	w	
Tue	5/23/2023	4.8	4.4	4.9	w	
Wed	5/24/2023	4.6	4.3	5	w	
Thu	5/25/2023	4.4	4.1	5.1	w	
Fri	5/26/2023	4.4	4.1	5.1	w	
Mon	5/29/2023	<hr/>				Closed Holiday
Tue	5/30/2023	4.6	4	5	w	
	5/31/2023	4.5	4.1	4.8	w	

CTSU Lab Refrigerator Min/Max Temperature Log

Maintain temperature between 2°C to 8 °C

Month **April**

Year **2023**

Refrigerator ID/Location:

Sudler

	Date (Weekdays only)	Time	Min °C	Max °C	Initials	Comments
Mon	4/3/2023	3.3	2.9	4.3	mw	
Tue	4/4/2023	3.4	3.2	4.2	PL	
Wed	4/5/2023	4	2.7	4.4	w	
Thu	4/6/2023	4	2.7	4.2	w	
Fri	4/7/2023	4.1	3.2	4.5	w	
Mon	4/10/2023	4.4	3.2	4.6	PL	
Tue	4/11/2023	4.2	3.2	4.6	w	
Wed	4/12/2023	4.4	4	4.6	w	
Thu	4/13/2023	4.4	4	4.6	w	
Fri	4/14/2023	4.2	3.9	4.6	w	
Mon	4/17/2023	4.8	3.8	5.0	OH	
Tue	4/18/2023	4.9	3.8	5.0	OH	
Wed	4/19/2023	4.4	3.8	5.0	PL	
Thu	4/20/2023	5.0	3.8	5.0	OH	
Fri	4/21/2023	4.6	3.8	5.0	w	
Mon	4/24/2023	4.1	3.9	4.6	mw	
Tue	4/25/2023	4.1	3.9	4.7	PL	
Wed	4/26/2023	4.1	3.8	4.7	PL	
Thu	4/27/2023	4.1	3.8	4.8	mw	
Fri	4/28/2023	4.1	3.8	4.8	w	

Lab Refrigerator Temperature Log for 2023

Maintain temperature between 1 - 10 C.

Fridge 1 to 10 °C

50
40
30
20
15
14
13
12
11
10
9
8
7
6
5
4
3
2
1
0
-1
-2
-3
-4
-5

	Jan	Feb	Min/ Max	Mar	Min/ Max	April	Min/ Max	May	Min/ Max	June	Min/ Max	COMMENTS
1	---	3.8 PL	3.4/4.2	3.6 PL	3.2/4.2	---	---	---	---	---	---	
2	---	4.0 PL	3.4/4.2	4.2 PL	3/4.3	---	---	---	---	---	---	
3	4.3 PL	3.7/4.5	3.4/4.2	4.2 PL	3/4.3	---	---	---	---	---	---	
4	3.7 PL	3.6/4.5	---	---	---	---	---	---	---	---	---	
5	3.7 PL	3.6/4.5	---	---	---	---	---	---	---	---	---	
6	3.7 PL	3.7/4.5	4.9 PL	3.4/3.9 PL	3.2/4.3	---	---	---	---	---	---	
7	---	---	4.1 PL	3.4/4.3 PL	3.2/4.3	---	---	---	---	---	---	
8	---	---	4.0 PL	3.4/4.3 PL	3.2/4.2	---	---	---	---	---	---	
9	4.2 PL	3.5/4.3	4.9 PL	3.2/5.1 PL	3.9 PL	3.1/4.2	---	---	---	---	---	
10	4.1 PL	3.4/4.4	5.1 PL	4.8/5.5	---	---	---	---	---	---	---	
11	4.2 PL	3.4/4.4	---	---	---	---	---	---	---	---	---	
12	3.8 PL	3.4/4.4	---	---	---	---	---	---	---	---	---	
13	3.8 PL	3.5/4.9	5.2 PL	5/5.6 PL	3.8 PL	2.9/4.5	---	---	---	---	---	
14	---	---	5.1 PL	5.1/5.5 PL	3.7 PL	2.9/4.5	---	---	---	---	---	
15	---	---	4.5 PL	5.1/5.5 PL	3.8 PL	2.9/4.5	---	---	---	---	---	
16	---	---	4.2 PL	5.1/5.5 PL	3.8 PL	2.9/4.5	---	---	---	---	---	
17	3.7 PL	3.7/11.5	4.7 PL	2.0/2.1 PL	4 PL	2.9/4.5	---	---	---	---	---	
18	4.0 PL	3.7/11.5	---	---	---	---	---	---	---	---	---	
19	4.8 PL	3.7/11.5	---	---	---	---	---	---	---	---	---	
20	4.8 PL	3.2/11.5	3.8 PL	3.5/5.3 PL	4.1 PL	3.6/4.2	---	---	---	---	---	
21	---	---	4.7 PL	3.7/4.2 PL	4.1 PL	3.3/4.2	---	---	---	---	---	
22	---	---	3.8 PL	3.5/4.3 PL	5.0 PL	3.3/4.2	---	---	---	---	---	
23	3.9 PL	3.7/11.5	3.6 PL	3.3/4.5 PL	4.9 PL	3.3/4.3	---	---	---	---	---	
24	3.8 PL	3.4/4.2	3.6 PL	3.3/4.3 PL	5.1 PL	3.3/4.3	---	---	---	---	---	
25	4.2 PL	3.4/4.5	---	---	---	---	---	---	---	---	---	
26	4.3 PL	4.1/4.5	---	---	---	---	---	---	---	---	---	
27	4.5 PL	4.1/4.5	4.2 PL	2.5/4.2 PL	5.0 PL	3.2/4.3	---	---	---	---	---	
28	---	---	4.0 PL	3.4/4.2 PL	3.8 PL	2.9/4.3	---	---	---	---	---	
29	---	---	---	---	3.4 PL	3.2/4.4	---	---	---	---	---	
30	3.7 PL	3.1/4.0	---	---	3.6 PL	2.8/4.4	---	---	---	---	---	
31	3.7 PL	3.1/4.0	---	---	3.6 PL	2.8/4.4	---	---	---	---	---	

-If out of range, recheck in 5 minutes.


-If remains out of range >5minutes, advise management.

Signatures: MW PH

Lab Refrigerator Temperature Log for 2022

Maintain temperature between 1 - 10 C.

Fridge
1 to 10
°C



	July	Aug	Sept	Oct	Nov	Dec	Min/Max	COMMENTS
1	4.9 DH 3.3/5.1	4.9 DH 3.0/5.7	4.9 W 3.5/4.5		4.3 W 3.5/5.7	3.9 W 3.5/4.5	3.5/4.5	
2		4.8 DH 3.0/5.7	5.1 W 3.3/4.1		4.2 PL 3.7/4.7	3.9 PL 3.5/4.5	3.5/4.5	
3		4.7 DH 3.0/5.7			3.9 W 3.4/4.0			
4		4.9 W 3.0/5.7			3.9 W 3.4/4.0			
5	5.0 DH 3.3/5.1	4.9 W 3.0/5.7			4.1 W 3.7/4.0	4.0 PL 3.4/4.5	3.4/4.5	
6	4.9 DH 3.3/5.1		5.1 W 3.3/4.0	4.1 W 3.7/4.0		4.2 PL 3.4/4.5	3.4/4.5	
7	4.9 DH 3.3/5.1		5.1 W 3.3/4.0	4.1 W 3.7/4.0	4.3 W 3.8/4.5	4.1 PL 3.4/4.5	3.4/4.5	
8	5.0 DH 3.3/5.1	4.9 W 3.0/5.7	5.2 DH 3.7/4.5		4.1 PL 3.5/4.5	4.1 W 3.4/4.5	3.4/4.5	
9		4.7 W 3.1/4.0	5.2 DH 3.0/4.5		4.1 PL 3.3/4.0	4.1 W 3.4/4.5	3.4/4.5	
10		5 W 3.4/4.0		4.5 W 3.0/4.0				
11	4.7 DH 3.3/5.1	5 W 3.4/4.0		4.5 W 3.0/4.0				
12	5.0 DH 3.3/5.1	5.1 W 3.1/4.0	4 W 3.7/4.5	4.5 W 3.0/4.0		4.1 PL 3.3/4.5	3.3/4.5	
13	4.7 DH 3.3/5.1		4.7 W 3.5/4.5	4.1 W 3.5/4.5		4.2 W 3.3/4.5	3.3/4.5	
14	4.7 DH 3.3/5.1		4.7 W 3.5/4.5	4.1 W 3.5/4.5	4.2 W 3.3/4.8	4.2 PL 3.3/4.4	3.3/4.4	
15	4.7 DH 3.3/5.1	5.1 DH 3.5/4.7	3.8 W 3.4/4.5		4.3 PL 3.3/4.0	4.4 W 3.6/4.4	3.6/4.4	
16		5.0 DH 3.2/4.7	4 W 3.4/5.1		4.0 PL 3.3/4.0	4.4 W 3.6/4.4	3.6/4.4	
17		5.1 DH 3.5/4.7		4 W 4.4/4.4	3.9 W 3.3/4.8			
18	5.0 DH 3.3/5.1	4.8 DH 3.5/4.7		4.1 W 4.1/4.4	3.9 W 3.3/4.8			
19	4.7 DH 3.3/5.1	4.9 W 3.2/4.7	4 W 3.2/4.5	4.2 W 3.0/4.4		4.1 PL 3.6/4.4	3.6/4.4	
20	4.7 W 3.2/5.0		4 W 3.0/4.5	4.2 W 3.0/4.4		4.2 PL 3.4/4.4	3.4/4.4	
21	4.7 W 3.2/5.0		4 W 3.2/4.5	4.2 W 3.0/4.4		3.9 W 3.9/4.8	3.9/4.8	
22	4.6 W 3.8/4.7	4.0 W 3.9/4.5	4.3 W 3.2/4.8		4.3 W 3.0/4.8			
23		4.6 W 3.5/4.5	4.3 W 3.2/4.8		4.3 W 3.0/4.8			
24		4.7 W 3.1/4.4		4.3 W 3.0/4.4				
25	4.9 DH 3.8/4.7	4.7 W 3.5/4.5		4.3 W 3.0/4.4				
26	4.9 DH 3.7/5.7	4.7 W 3.5/4.5	4.5 W 3.2/4.0	4.3 W 3.0/4.4				
27	5.0 DH 3.7/5.7		4 W 3.0/4.0	4.3 W 3.3/4.4				
28	5.0 DH 3.7/5.7		4.3 W 3.9/4.5	4.3 W 3.5/5.7	4.2 W 3.8/4.4			
29	4.9 DH 3.7/5.7	4.7 W 3.3/4.5	4 W 3.4/4.0		3.8 PL 3.5/4.5			
30		4.9 W 3.5/4.9	4 W 3.4/4.0		4.2 PL 3.5/4.5			
31		4.9 W 3.5/4.5		4.3 W 3.5/5.7				


-If out of range, recheck in 5 minutes.
-If remains out of range >5 minutes, advise management.

Signatures:  

Lab Refrigerator Temperature Log for 2022

Maintain temperature between 1 - 10 C.

Fridge
1 to 10
°C



	Jan	Feb	Mar	April	May	June	Min/Max	Comments
1				5.3m		5.1DH	3.5/5.1	
2					5.3m	5.7DH	3.5/5.1	
3					5.4DH	5.5DH	3.8/5.1	
4				5.2m	5.4m	5.8DH	3.8/5.1	
5				5.2m	5.4m			ERR 06/03/2022
6				5.0DH	5.3DH	5.3DH	3.8/5.1	
7				5.1DH		5.3DH	3.7/5.1	
8				5.1DH		5.5DH	3.7/5.1	
9					5.4DH	5.2DH	3.7/5.1	
10					5.3m	5.4DH	3.7/5.1	
11				5.0m	5.3m			
12				5.1DH	5.4m			
13				4.9DH	5.5m	5.4DH	3.7/5.1	
14				5.1m		5.1DH	3.4/5.1	
15			4.9m	5.2DH		5.8DH	3.6/5.1	
16			4.8DH		5.0DH	5.2DH	3.5/5.1	
17			4.9DH		5.2DH	5.5DH	3.5/5.1	
18			5.0DH	5.1DH	5.2m	5.8DH	3.5/5.1	
19				5.0DH	5.2m	5.2DH	3.3/5.1	
20				5.0DH	5.5m	5.1DH	3.3/5.1	
21			5.1m	5.0DH		4.8DH	3.3/5.1	
22			5.1m	4.9DH		4.8DH	3.3/5.1	
23			5.1m		5.5m			
24			4.9m		5.4m			
25			5.1m	5.0DH	5.4DH	4.9DH	3.3/5.1	
26				5.1m	5.3DH	5.0DH	3.3/5.1	
27				5.1m	5.3DH	5.4DH	3.4/5.1	
28			4.9m	4.9m		5.4DH	3.4/5.1	
29			5.0DH	5.4DH		5.3DH	3.4/5.1	
30			5.0m			4.9DH	3.7/5.1	
31			5.3m		5.2DH			

Signatures: MW

DH Page RW

-if out of range, recheck in 5 minutes.

-if remains out of range >5minutes, advise management.


Lab Refrigerator Temperature Log for 2022

Maintain temperature between 1 - 10 C.



Fridge

1 to 10 °C

°F °C



	Jan	Feb	Mar	April	May	June	July	Aug	Sept	Oct	Nov	Dec
1	5.0m	4.5m										
2	closed	4.7m										
3	5.1m	closed	4.8m									
4	4.7m	4.9m	5.5m									
5	4.8m											
6	4.8m											
7	4.8m	4.9m	4.8m									
8		5.0m	4.6m									
9		5.0m	4.9m									
10	4.8m	5.2m	4.9m									
11	4.9m	5.2m	4.9m									
12	5.1m											
13	4.7m											
14	5.1m	5.1m	5.1m									
15		5.3m										
16		5.4m										
17		closed										
18	4.9m	5.3m										
19	4.8m											
20	4.8m											
21	4.9m	5.1m										
22		4.7m										
23		5.1m										
24	5.0m	5.1m										
25	4.9m	5.1m										
26	5.0m											
27	4.7m											
28	4.8m	5.0m										
29												
30												
31	4.7m											

Signatures: MW  

-if out of range, recheck in 5 minutes.
 -if remains out of range >5minutes, advise management.

Lab Specimen Refrigerator Temperature Log for 2021


Fridge

to 10 °C

°F °C

Maintain temperature between 1 - 10 C.

	Jan	Feb	Mar	April	May	Jun	July	Aug	Sept	Oct	Nov	Dec
1	4.8m	5.0m	5.1m	4.2m	—	5.1m	4.4m	—	4.5m	4.4m	5.3m	5.2m
2	4.8m	5.0m	5.1m	4.4m	—	5.0m	4.5m	4.8m	4.4m	—	5.1m	5.2m
3	4.8m	5.1m	5.2m	—	4.7m	5.4m	—	4.6m	—	4.5m	5.2m	—
4	4.8m	5.2m	5.3m	4.4m	5.0m	5.0m	—	4.6m	—	4.7m	5.0m	—
5	4.9m	—	—	4.5m	5.1m	—	4.7m	4.2m	—	4.3m	—	5.0m
6	4.9m	—	—	4.6m	5.2m	4.9m	4.5m	—	4.5m	4.5m	—	5.1m
7	4.7m	5.1m	5.0m	4.6m	—	5.0m	4.6m	—	4.4m	5.1m	5.0m	5.0m
8	—	4.8m	5.1m	4.6m	—	5.0m	4.6m	—	4.4m	—	5.3m	5.2m
9	—	4.9m	5.3m	—	5.0m	5.3m	—	5.5m	4.8m	—	5.0m	4.9m
10	4.8m	4.9m	5.0m	—	5.0m	5.2m	—	4.4m	—	4.6m	—	—
11	4.9m	4.7m	5.3m	4.3m	5.3m	—	4.4m	4.7m	—	4.6m	5.2m	—
12	5.0m	—	—	4.4m	5.2m	—	4.3m	4.7m	4.8m	4.4m	—	4.7m
13	4.8m	—	—	4.3m	5.2m	—	5.0m	—	4.1m	4.6m	—	5.3m
14	4.8m	5.0m	5.0m	4.8m	5.0m	5.1m	5.0m	—	4.7m	4.9m	5.1m	5.2m
15	—	4.8m	5.0m	4.8m	—	5.3m	4.8m	4.7m	4.2m	4.9m	5.0m	5.2m
16	—	5.0m	5.0m	—	5.0m	4.9m	—	4.4m	4.8m	—	5.0m	5.4m
17	—	4.8m	4.4m	—	5.0m	5.0m	—	4.5m	—	5.1m	5.2m	—
18	5.0m	4.8m	4.4m	4.8m	5.2m	—	4.8m	4.3m	—	5.2m	5.1m	—
19	4.9m	—	—	4.9m	5.0m	—	4.7m	4.5m	4.5m	5.4m	—	5.3m
20	4.9m	5.0m	4.4m	5.0m	4.8m	5.2m	4.8m	—	4.5m	5.1m	—	4.5m
21	4.9m	—	—	5.0m	—	4.5m	4.7m	—	4.7m	5.1m	5.1m	5.0m
22	—	5.0m	4.6m	5.0m	—	4.5m	4.8m	4.5m	4.6m	—	5.1m	5.0m
23	—	5.1m	4.5m	—	5.0m	4.2m	—	4.4m	4.6m	—	5.0m	—
24	5.1m	4.9m	4.2m	—	4.9m	4.4m	—	4.4m	—	4.8m	—	—
25	4.7m	5.0m	4.4m	5.0m	4.9m	—	4.9m	4.7m	—	5.4m	—	—
26	4.9m	—	—	5.0m	4.9m	—	4.9m	4.7m	4.5m	5.3m	—	—
27	4.8m	—	—	4.9m	4.8m	4.3m	4.7m	—	4.5m	5.4m	—	—
28	4.9m	—	—	5.0m	—	4.3m	—	—	4.4m	5.3m	5.0m	5.0m
29	—	—	4.6m	5.0m	—	4.3m	4.6m	—	4.5m	5.3m	5.0m	5.0m
30	—	—	4.5m	4.7m	—	4.5m	4.4m	4.4m	4.5m	—	5.1m	—
31	—	—	4.7m	—	—	—	—	4.5m	—	—	—	—



-if out of range, recheck in 5 minutes.
 -if remains out of range >5minutes, advise management.

Signatures: MW JPB
 JPB

