

Service Report



WORK ORDER ID: CWKD6975329

WORK ORDER TYPE: Preventative Maintenance

CUSTOMER: UNIVERSITY OF KS MED CTR-CTSU | **DEPARTMENT:** 146148-UNIVERSITY OF KS MED CTR-CTSU-BIOMED | **ACCOUNT NUMBER:** 146148

CONTACT:

CUSTOMER PO#:

CONTACT PHONE:

CUSTOMER REFERENCE:

BILLING ADDRESS: PO Box 3160 Mail Stop 6011 Kansas City, KS 66103

SHIPPING ADDRESS: 4350 Shawnee Mission Parkway Mail Stop 6011 Fairway, KS 66205

AGILITI ASSET ID : 1485862 | **SERIAL #:** ALZ19A019 | **VENDOR SITE ID:** | **CUSTOMER ASSET ID:**

ASSET TYPE: CENTRIFUGES, TABLETOP, HIGH-SPEED, REFRIGERATED | **MANUFACTURER:** BECKMAN COULTER INC | **MODEL:** ALLEGRA X-30R

SERVICE REQUESTED: PM DUE

SERVICE PROVIDED TYPE: Scheduled Maintenance

WORK COMPLETED: 01-22-2025

LABOR

SERVICE PROVIDER	CATEGORY	HOURS	PERIOD DEFINITION	SERVICE DATE
Stephanie Dye	PM Not Covered	0.4	M-F 8-5	01/22/25

SERVICE PROVIDED DETAILS

UNIT PASSED ALL REQUIRED TESTING AT THIS TIME

TEST EQUIPMENT

PREFIX	UNIT	DESCRIPTION	CALIBRATION DATE
TA1	0257	TACHOMETER	04/05/24

CHECKLIST

PROCEDURE: CENTRIFUGE

QUESTION

BRUSHES
COVER LATCH
Electrical Safety Inspection
Mechanical Safety Inspection
MOTOR CLEANED
MOTOR CURRENT
ROTOR & MOTOR
RPM CHECK
SAFETY INTERLOCK
TACHOMETER SERIAL NUMBER
TEMPERATURE VERIFICATION
TIMER

ANSWER

Pass
Pass
Pass
Pass
Yes
Pass
Pass
Pass
Pass
TA10257
Pass
5 MIN

Customer Signature:

Stephanie Dye

Technician Signature:

Date:

1/22/25

Date: