



**Blood Presure Monitor Performance Report**

Customer:		KU-CTSU Swope				Contact:		Autumn Beauchamp									
Street:		4001 Dr. Martin Luther King jr Blvd				Phone #		913-574-4594									
City/ST/Zip:		Kansas City, MO 64130				P= Pass F=Fail											
MFG	Model#	S/N	Electrical Safety Check	Condition Controls/ Switches	Condition Power Cord & Strin Relief	Hardware Integrity	Indecators & Display Operation	MFG. Labeling Intact	Normal		Hypertonic		Hypotonic		Oral Themp (F)	SPO2	Date Performed
									120/80	P 60	200/150	P130	60/30	P40	98.6	97%	

**Blood Presure Monitor Performance Report**

Customer:		KU-CTSU Swope			Contact:		Autumn Beauchamp										
Street:		4001 Dr. Martin Luther King jr Blvd			Phone #		913-574-4594										
City/ST/Zip:		Kansas City, MO 64130			P= Pass F=Fail												
MFG	Model#	S/N	Electrical Safety Check	Condition Controls/ Switches	Condition Power Cord & Strin Relief	Hardware Integrity	Indecators & Display Operation	MFG. Labeling Intact	Normal		Hypertonic		Hypotonic		Oral Themp (F)	SPO2	Date Performed
									120/80	P 60	200/150	P130	60/30	P40	98.6	97%	