

CTSU Lab Freezer Min/Max Temperature Log

Maintain temperature between -65°C to -90°C

Month **February**

Year **2024**

Freezer ID/Location: Lab room freezer 841803-1140 (-80 freezer A)

	Date (Weekdays only)	Time	Min °C	Max °C	Initials	Comments
Thu	2/1/2024	0900	-87.98	-55.34	DW	%
Fri	2/2/2024	0900	-87.44	-55.34	DW	%
Mon	2/5/2024	0800	-86.59	-77.83	DW	
Tue	2/6/2024	0800	-86.52	-68.86	LM	
Wed	2/7/2024	0800	-86.89	-77.68	DW	
Thu	2/8/2024	0800	-87.06	-62.40	DW	#
Fri	2/9/2024	0800	-87.28	-67.06	DW	
Mon	2/12/2024	0900	-87.41	-78.16	DW	
Tue	2/13/2024	0900	-87.3	-68.77	DW	
Wed	2/14/2024	0800	-87.31	-66.52	LM	
Thu	2/15/2024	0900	-87.40	-69.72	DW	
Fri	2/16/2024	0800	-87.72	-62.12	LM	
Mon	2/19/2024	1000	-87.57	-61.43	DW	#
Tue	2/20/2024	0915	-87.67	-61.43	DW	%
Wed	2/21/2024	0900	-87.67	-62.85	DW	%
Thu	2/22/2024	1000	-87.59	-75.81	DW	
Fri	2/23/2024	1000	-87.69	-76.77	DW	
Mon	2/26/2024	0800	-87.76	-78.12	DW	
Tue	2/27/2024	0800	-87.64	-71.15	DW	
Wed	2/28/2024	1000	-87.64	-71.15	DW	
Thu	2/29/2024	1000	-87.34	-65.27	DW	

Key: * Heavy usage
 % Normal peak of cooling curve
 # Normal trough of cooling curve